A-A-53092A
31 July 1990
SUPERSEDING
A-A-53092
31 August 1987

COMMERCIAL ITEM DESCRIPTION

TRAY, INSTRUMENT

(Corrosion-Resisting Steel)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification RR-T-636 which is cancelled.

This Commercial Item Description covers a corrosion-resisting steel instrument tray.

Salient characteristics:

The instrument tray shall be rectangular and shall be supplied in the following nominal sizes:

Size 1 - 8-7/8 inches long by 5 inches wide by 2 inches deep, with cover.

Size 2 - 10-1/2 inches long by 8 inches wide by 2 inches deep. As an alternate, 10 by 8-1/4 by 2-1/2 inches.

Size 3 - 15-1/2 inches long by 9-1/2 inches wide by 2 inches deep.

Size 4 - 19-1/4 inches long by 12-3/4 inches wide by 3/4 inch deep.

Size 5 - 20-5/8 inches long by 12-5/8 inches wide by 4 inches deep.

Size 6 - 16-3/4 inches long by 9-1/4 inches wide by 2-1/2 inches deep, with perforated bottom.

Size 7 - 15-1/8 inches long by 10-1/2 inches wide by 5/8 inch deep.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Complete dimensions for the finished trays are shown on the following Defense Personnel Support Center (DPSC) drawings:

Drawing 20686	Tray size 1
Drawing 20460	Tray sizes 2, 3, 5, and 7
Drawing 20019	Tray size 4
Drawing 24019	Tray size 6

The body of the tray shall be seamless die formed from a single piece of corrosion-resising steel, American Iron and Steel Institute (AISI) type 302or 304, and shall rest evenly on a flat surface without rock or wobble. In overall dimensions, the tray shall be square within plus or minus 1/8 inch. The starting thicknesse of the metal used and the minimum final thicknesses afte forming shall be as shown below:

Size	Original Metal Thickess	Minimum Final Thickness
1	0.031 inch	0.024 inch
2	0.031 inch	0.027 inch
3	0.031 inch	0.024 inch
4	0.031 inch	0.020 inch
5	0.037 inch	0.024 inch
6	0.031 inch	0.020 inch
7	0.031 inch	0.020 inch

The bottom of the tray may cave inward so that the highest point shall be no more than 6 inch above the surface upon which it rests.

The cover for the size 1 tra hall be seamlessly die formed from a single piece of corron-resisting steel, AISI type 302 or 304, not less than 0. inch thick, which shall not be reduced below 0.0[illegible] inch at [illegible] point in the finished cover. The design and dimensions sh be as indicated on DPSC Drawing 20686. The rim of t cover shall be formed to fit snugly over the bead on the tom of the tray, and shall have

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

a lift handle of the design shown on DPSC Drawing 20686. The handle shall be made of corrosion-resisting steel, AISI type 302 or 304, not less than 0.044 inch thick, which shall not be reduced below 0.038 inch at any point in the finished handle. The lift handle shall be centered within 1/16 inch and spot-welded to the top of the cover.

The cover may be of either design as shown on DPSC Drawing 20686. In overall dimensions, the cover and handle shall be square within plus or minus 1/8 inch. The cover shall not rock on the body of the tray.

The perforations in the size 6 instrument tray shall be as shown on DPSC Drawing 24019.

The entire inside and outside surfaces of the tray and cover shall be finished to a degree that will facilitate cleaning, such as a natural scratch-free mill finish or better. A scratch mark that can be felt with a fingernail shall be unacceptable. Grinding shall produce a uniform finished surface with all imperfections and marks of previous grinds removed. The thickness of the sheet forming the tray or cover shall not be reduced below the minimum thickness specified by grinding to remove surface imperfections. As an alternate, the tray and cover shall be finished by electropolishing. The electropolishing shall result in a uniform, bright, and shiny surface which can be readily cleaned. As a final operation, the tray and cover shall be thoroughly cleaned to remove foreign particles such as dirt, grease, graphite, and metal.

The tray shall be free of rough grind marks, trimming marks, pits, scales, and other defects or imperfections. All surfaces shall have a smooth finish and shall be free of burrs or sharp edges.

Workmanship. The instrument tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tray of the size specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one instrument tray.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one instrument tray.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-00-793-9600 - (Size 1)

6530-00-793-9945 - (Size 2)

6530-00-794-0000 - (Size 3)

6530-00-793-9570 - (Size 4)

6530-00-794-0300 - (Size 5)

6530-00-914-0238 - (Size 6)

6530-00-926-9170 - (Size 7)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6530-2132

Location:
CID/TRAY/INSTRUMENT/H-13-37

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53092A	TRAY, INSTRUMENT, (Corrosion-Resisting Steel)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 8 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.