A-A-53097
31 August 1987

COMMERCIAL ITEM DESCRIPTION

RING, CERVICAL SPINE TRACTION

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a cervical spine traction halo ring suitable for use in the immobilization of cervical spine fractures.

Salient characteristics:

The cervical spine traction halo ring shall be specifically designed for immobilization of cervical spine fractures and shall be furnished in the following head circumference sizes:

Size 1	54 to 56 centimeters
Size 2	54 to 58 centimeters
Size 3	58 to 60 centimeters

The ring shall be fabricated of a suitable corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series, properly tempered and passivated.

The lateral sides of the ring shall have many holes through which skull positioning pins and skull halo pins are inserted. The location of the holes on one side of the halo ring shall be spaced identically to those on the opposite side.

One hole shall be located at the central part of the rounded frontal portion of the halo. This hole, when placed over the bridge of the nose, is a guide for proper halo alignment.

The occipital portion of halo shall form an upward curve to allow for high cervical surgery.

The ring shall have a #4 commercial finish, or better, and be readily cleanable.

The ring shall be free of cracks, sharp edges, burrs, pits, dents, ripples, and other deformities.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Each ring shall be furnished with one copy of comprehensible instructions for the assembly and use of the item and possible adverse effects that may result. The instructions shall include a picture of the ring assembled and in place on a patient and shall include instructions for cleaning. The contractor shall be identified in the instructions.

Workmanship. The cervical spine traction ring shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One halo ring, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one cervical spine traction halo ring as specified in this description.

TABLE I. Sampling for examination and tests.

	Inspection level	AQL (defects per hundred units)
VISUAL EXAMINATION	II	
Major		2.5
Total (Major and minor)		6.5
DIMENSIONAL EXAMINATION	S-2	2.5
TESTS	S-2	Acceptance number is zero

Examination. The halo ring shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Classification of Defects

Classification	Defect
Major	
101	Missing, improper alignment of, or improperly located components.
102	Not free from construction defects which weaken the strength or durability of the item for its intended use or cause discomfort to the user.
103	Not free from loose components or improperly applied fastening, or fasteners.
104	Components or assembly not as described in the salient characteristics.
105	Not free from burrs and sharp edges.
Minor	
201	Instructions not furnished or not complete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Dimensional Examination. The halo ring shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Tests. Testing shall be conducted to determine compliance with the salient characteristics specified herein. Failure of any test shall be considered a major defect. Testing shall include but not be limited to the following:

The halo ring shall be assembled and mounted on an individual or mannequin of the appropriate size to match the size of the halo ring under test. Halo ring shall allow for 1-1/2 cm. clearance all around the head. The test unit shall be examined for compatibility with one of the traction apparatuses noted on page 5. Noncompliance shall be cause for rejection.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6515-01-166-7846 (Size 1)
6515-01-166-7782 (Size 2)
6515-01-168-0557 (Size 3)

Note: The Halo rings shall be designed for use with traction apparatus, NSN 6515-01-166-7781, NSN 6515-01-153-1676, and NSN 6515-01-153-1675.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4014

PGC 06211

Location: CID/RING/H13-24

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 8, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53097	Ring, Cervical Spine Traction

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

TO DETACH THIS FORM, CUT ALONG THIS LINE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.