A-A-53098
31 August 1987

COMMERCIAL ITEM DESCRIPTION

CALIBRATOR, TIMER

The General Services Administration has approved the use of this Commercial Item Description.

This Commercial Item Description covers a portable, battery-operated, digital, x-ray generator timer.

Salient Characteristics:

The x-ray generator timer shall be non-invasive, portable, battery-operated, and digital. The timer shall be capable of measuring the timer accuracy of half-wave, full-wave, three phase, and constant-potential generators.

The timer shall be capable of displaying measurements within the following three measuring ranges:

Pulse Range:	0-19,999 Pulses
Timer Range:	0-1999.9 MS
	0-19.999 Sec.

The timer accuracy shall be ± 0.1 MS, ± 1 MS, or ± 1 pulse, depending on range selection.

The digital display shall be easy-to-read, approximately 1/2 inch high. The display shall remain in view until the next exposure and shall be capable of resetting automatically with each exposure.

The timer shall have a visible or audible over-range indicator and low battery indicator.

The timer shall have a power on/off switch and a 3-position range selector switch.

AMSC/NA FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The timer shall have the following approximate dimensions:

Height:	2-1/2 inches (6.4 cm)
Width:	6 inches (15.2 cm)
Depth:	6-1/4 inches (15.9 cm)

The timer shall have an output connector to allow analysis of the radiation waveform using a storage oscilloscope.

The battery shall be easily replaced, and provide at least 100 hours of continuous use.

The timer finish shall be uniform and free of scratches, discoloration, and burrs.

Each timer shall be supplied with a set of operating instructions, and service manual with parts identification.

Workmanship. The timer shall be free from defects wich detract from its appearance or impair its serviceability.

Unit. Each (EA). One timer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the currenversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for test purposes shall be one timer calibrator.

TABLE I. Sampling for testing and examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5
For All Tests	S-1	1.5

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Each timer calibrator shall be examined to detrmine compliance with the classification of defects in table II.

TABLE II. Classification of Defects *

Classification	Defects
Major	
101	Pulse display does not function.
102	Timer display does not function.
Minor	
201	Identification marking not complete, permanent or legible.
202	Case not free of chips, cracks, scratches, and other appearance defects.
203	Any component part not secured in place.

*Examination shall not be limited to the classified possible defects listed herein.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. All tests shall be performed on a first article sample in accordance with table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

TABLE III. Test.

1.	Unit shall demonstrate capability of displaying pulses from 0 to 19,999.
2.	Unit shall demonstrate capability of displaying pulses from 0 to 1999.9 mS, and 0 to 19.999 seconds.
3.	Unit shall demonstrate capability of resetting automatically with each exposure.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6625-01-185-8448

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 52320

Project No. 6625-0661

Location: CID/CAL/H13-12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4986 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A_53098	Calibrator, Timer

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.