A-A-53100
31 August 1987

COMMERCIAL ITEM DESCRIPTION

GLOVES, X-RAY PROTECTIVE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-G-36479, which is cancelled.

This Commercial Item Description covers x-ray protective gloves, suitable for protecting hands and forearms of wearer from exposure to x-rays.

Salient characteristics:

The x-ray protective gloves shall be a pair of lead impregnated vinyl plastic gloves with five fingers, soft absorbent cotton lining, and two pairs of slip-on (removable) external cotton sheaths with stretch cuffs.

The gloves shall be medium size (commercial size 10) and not less than 15 inches (38.1 cm) in overall length.

The protective elements of each glove shall be highly flexible leaded vinyl molded in one piece with no seams, and shall provide the wearer protection equivalent to not less than 0.5 mm lead at 80 PKV, except that any individual area smaller than 2 mm square or a summation of all such areas amounting to not more than 12 square mm may have a lead equivalent below 0.5 mm, but not less than 0.25 mm.

The gloves shall be unaffected by temperatures ranging from -30°F to +140°F. They shall be sufficiently flexible to permit operation of x-ray apparatus controls, palpation of patients, and be capable of withstanding continuous hard use and repeated flexing without cracking.

The cotton sheaths supplied (two per glove) shall be made from a suitable washable, color-fast, and durable, cotton material. The sheaths shall be long enough to extend over the ends of the vinyl glove elements and be held taut by the stretch cuff.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The gloves shall be supplied as a matched pair, one for right hand and one for left hand usage. The gloves shall be free of cracks, cuts, tears, holes, perforations, or blisters.

Workmanship. The x-ray protective gloves shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Pair (PR). One pair of x-ray protective gloves, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one pair of gloves as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The gloves shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Classification of Defects

Classification	Defects
Major	
101	Gloves not sufficiently flexible to permit operation of x-ray controls.
102	Gloves not free of cracks, cuts, tears, holes, perforations, or blisters.
103	Gloves not a matched pair.
104	Cotton glove inserts or sheaths not supplied.
105	Size not as specified.
Minor	
201	Gloves not clean and free of foreign material.

Dimensional Examination. The gloves shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Tests.

Lead Equivalent test.

Two samples of each glove shall be radiographed at 80 KVP with intensifying screens, at an appropriate milliampere-seconds and target-film distance, using a lead step wedge. The lead equivalent of the samples shall then be determined. Failure to achieve an equivalency above the specification levels of 0.5 mm and 0.25 mm on any of the samples shall constitute a major defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Temperature tests.

The gloves shall be subjected to a high temperature test and a low temperature test. Tolerance on each temperature shall be ±5°F. At the end of each test, the gloves shall be put on, repeatedly flexed, and examined for cracks, breaks, and deterioration.

Low temperature - The low temperature test shall be -30°F for 12 hours followed by an ambient of 72°F for a minimum of 6 hours.

High temperature - The high temperature test shall be at +140°F for 72 hours followed by an ambient of 72°F for a minimum of 6 hours.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6532-00-926-8920

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6532-1176

PGC 50905

Location: CID/GLOVES/H13-23

☆U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-036-75870

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.