COMMERCIAL ITEM DESCRIPTION

APPLICATOR, DISPOSABLE, (Wood or Plastic, Cotton or Polyester Tipped End)

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers a sterilized or sterilizable fiber tipped wooden or plastic applicator, for use in the application of medicaments or biological cultures.

Salient Characteristics.

Shall be a cotton or polyester tipped applicator for use in the application of medicaments in a dispensary or biological cultures in a laboratory.

Shall be supplied in a sterilized or a sterilizable condition.

Shall be supplied with a wooden or plastic applicator stick having nominal dimensions (inches) of 1/12 diameter by 6 length, with commercial tolerances acceptable. The applicator stick shall be symmetrical throughout its length and its ends shall be cut square and smooth.

Shall have fibers affixed to one end of the applicator stick by use of a nontoxic adhesive that shall inhibit mold growth, and is capable of withstanding sterilization without affecting the intended use of the tipped applicator.

Shall have fibers for applicator stick fabricated of either:

a. Purified cotton conforming to requirements of the United States Pharmocopeia (USP).
b. 100% polyester fiber.
c. A polyester/cotton blend.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Shall have a wooden applicator stick, if supplied, fabricated of tough, flexible, well seasoned, fungus-free wood that is smoothly finished, and has a moisture content of not less than 6% and not more than 12%.

Shall have a plastic applicator stick, if supplied, fabricated of a suitable plastic material which has sufficient flexibility to preclude breaking when bent to an angle of at least 80 degrees at its midpoint.

The method and efficacy of sterilization, for an applicator supplied in a sterile condition, shall be acceptable to the Food and Drug Administration (FDA).

The sterilizable (non-sterile) applicator, if supplied, shall be provided together with other applicators, constituting the same unit of issue quantity, in a suitable package that will permit and maintain sterilization of the applicator contents. This package shall contain a sterilization indicator for the required method of sterilization (gas or steam). This package shall also be supplied in a sealed condition that can withstand sterilization and maintain sterility of its applicator contents until the package is opened.

Workmanship. The applicator and its immediate package shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing one hundred sterile or sterilizable applicators, as specified, constitutes one unit. The sterile applicators shall be packaged in a sealed container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and Tests. Inspections and Tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6515-01-234-6838

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

PROJECT NO. 6515-3943

Location: NCR Disk # 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53101	APPLICATOR, DISPOSABLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.