A-A-53106
31 August 1987

COMMERCIAL ITEM DESCRIPTION

APPLICATOR, CRANIAL IMMOBILIZATION RING

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a halo ring application device for simplifying the application of cranial halo rings.

Salient characteristics:

The cranial immobilization ring applicator shall be a device for simplifying the application of cranial halo rings which are used for immobilization of cervical spine fractures. Secured by the patient's own weight, the halo ring application device shall safely support both the patient's head and the halo rings, leaving the surgeon's hands free for skull pin insertion. The device shall significantly increase the speed, accuracy, and safety of halo application.

The applicator device shall consist of a strong, high impact base shaped like an ironing board, an integral swivel block assembly at the squared end of the base, and a head support rod assembly.

The swivel block assembly, which shall be an integral part of the high impact base, shall have two holes to securely seat the two longer rods of the head support rod assembly. The sides of the swivel block assembly shall be perforated to securely accommodate a strong metal adjustment pin. Head support and rod angulations from 0 degrees to 60 degrees shall be obtained through the use of the adjustment pin.

The head support rod assembly shall consist of two long round rods with an integral cross-bar at one end extended to form a "T" at one side of the assembly, and a separate adjustable rod with a universal joint assembly. The two longer rods shall fit securely into the holes in the swivel block assembly. The universal joint assembly shall slide into the extended crossbar and shall accommodate the separate adjustable rod. The halo rings shall attach into the adjustable rod and be secured with a locking screw-type mechanism.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The metal portions of the cranial immobilization ring applicator shall be fabricated of a suitable corrosion-resisting steel, American Iron and Steel Institute, (AISI) 300 series, properly tempered and passivated.

All welding shall be durable, sound in every detail, free of cracks and porosity, and in accordance with the latest practices recommended by the American Welding Society.

The cranial immobilization ring applicator shall be furnished thoroughly clean and free of dirt, grease, metal particles, and other foreign material. The applicator shall be capable of being easily cleaned between uses.

Each applicator device shall be furnished with one copy of easily understood instructions for the assembly and use of the device and possible adverse effects that may result. The instructions shall include a picture of the device assembled and in place on a patient and shall include instructions for cleaning. The contractor shall be identified in the instructions.

Workmanship. The cranial immobilization ring applicator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cranial immobilization ring applicator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one cranial immobilization ring applicator as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The cranial immobilization ring applicator shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Classification of Defects.

Classification	Defect
Major	
101	Any component part missing, broken, or otherwise unusable.
102	Metal portions of applicator not corrosion-resisting steel.
103	Head support and rod angulations from 0 degrees to 60 degrees not obtainable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

TABLE II. Classification of Defects - Continued.

Classification	Defect
Major cont'd	
104	Applicator not capable of being easily cleaned between uses.
105	Instructions for assembly and use not included.
106	Welding not durable, not free of cracks and porosity.
Minor	
201	Applicator not furnished thoroughly clean and free of dirt, grease, metal particles, and other foreign material.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-166-5068

The cranial immobilization ring applicator shall be for use with cervical immobilization traction apparatus, NSN's 6515-01-153-1676, 6515-01-153-1675, and 6515-01-166-7781.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3898

PGC 53048

Location: CID/APPLICATOR/H13-19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 8, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4986 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53106	Applicator, Cranial Immobilization Ring

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.