GG-C-00289a(DSA-DM)
September 11, 1969
Interim Revision of
GG-C-289
October 25, 1949

INTERIM FEDERAL SPECIFICATION

CHISELS, DENTURE-TRIMMING

This Interim Federal Specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this Interim Federal Specification as a valid exception to Federal Specification GG-C-289.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers chisels suitable for dental laboratory use in trimming dentures.

1.2 Classification. The chisels shall be of the following shapes, as specified (see 6.1):

Shape 1. Left bevel, No. 17.
Shape 2. Right bevel, No. 16.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following specifications, standards, and drawing, of the issue in effect on date of invitations for bid or requests for proposal, form a part of this specification to the extent specified herein:

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

STANDARDS

Federal

FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Fort Worth, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-P-116 Methods of Preservation.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC).

23204 Chisels, Denture Trimming, Laboratory.

(Copies of military specification, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitations for bid or requests for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM Designation E18 Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

ASTM Designation E92 Tests for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Handle. The handle of the chisel shall be fabricated from suitable, close-grained hardwood.

3.1.2 Blade and shank. The blade and shank of the chisel shall be fabricated from class 1, type 1095 carbon steel conforming to the chemical requirements of GG-I-526, except that no plating will be required and the corrosion tests shall not apply.

3.1.3 Ferrule. The ferrule shall be fabricated from either nickel-plated steel or chromium-over-nickel plated steel.

3.2 Style, design, and dimensions. The chisels shall be of the Pearson design, one with a left bevel (No. 17) and the other with a right bevel (No. 16). The chisels shall be suitable for use in the dental laboratory during the trimming of dentures. The style, design, and dimensions of the chisels shall be in accordance with DPSC Drawing 23204.

3.3 Construction. The chisels shall be of three-piece construction, consisting of a handle, a blade with shank, and a ferrule. The shank shall be securely attached to, and imbedded in, the handle. The imbedded portion shall be flattened for a minimum distance of 1/4 inch to prevent the shank from turning within the handle. The ferrule shall be securely attached to the distal end of the handle by crimping or by force fitting.

3.4 Hardness. After proper stress relieving, that portion of the blade immediately adjacent to the beveled edge shall have a Rockwell hardness of C50 to C61, when tested as specified in 4.4.1.

3.5 Finish. Operative ends (blades) of the chisels, including the ground and sharpened edges of the beveled portion, shall have a fine honed finish. The shanks and ferrules shall have a smooth and polished finish. The handle shall have a smooth, painted or lacquered finish, except that the underside of the knob portion of the handle may be free from lacquer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.6 Identification marking. The blade or shank of each chisel shall be permanently and legibly marked with the instrument number and the name or trademark of the manufacturer. Instruments of foreign (other than U. S.) manufacture shall also be marked with the name of the country of origin.

3.7 Workmanship. Workmanship shall be first class throughout. The denture trimming chisels shall be free from defects which detract from their appearance or may impair their serviceability, proper functioning, or intended use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the steel and wood, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables I and III (see 4.2.1 and 4.3) and table II (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I. The unit of product for examination purposes shall be one chisel.

Table I. Examination criteria

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for inspection purposes shall be one chisel. Acceptance number shall be zero for all sample sizes.

Table II. Test Criteria

Characteristic	Requirement	Test Procedure	Inspection level
Hardness	3.4	4.4.1	S-2
Chemical composition	3.1	4.4.2	/1

/1 A sufficient quantity of each lot of raw metal used in the construction of the blade and shank shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.3 Examination. The chisels shall be examined to determine compliance with the requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table III.

Table III. Classification of Defects

Categories	Defects
Major A:	
101	Style and design for the applicable shape chisel not as specified (see 3.2).
102	Cutting edge of operative end not sharp.
103	Chisel working surface not free from feathers, cracks, or burrs.
104	Operative end finish not as specified.
105	Construction not as specified.
Major B:	
151	Operative end dimensions not as specified.
Minor A:	
201	Shank, ferrule, or handle finish not as specified.
202	Chisel not free from dirt, polishing compound, or foreign matter.
203	Identification marking not present, improperly located, illegible, or incomplete.
204	Packaging and packing not as specified.
205	Marking on packaging and pack not as specified.
Minor B:	
251	Dimensions (other than 151) not within tolerances shown on drawing.

4.4 Test procedures.

4.4.1 Hardness. The hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.4.2 Chemical composition. In the absence of certificates of quality for the raw material, the material used to fabricate the blade and shank shall be subjected to chemical analysis in accordance with Method 111 or 112 of FED-STD-151. In case of dispute, Method 111 shall be used for referee analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A, or C, as specified (see 6.1).

5.1.1 Unit of issue. One chisel, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design, constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flaps with either adhesive (conforming to MMM-A-250) or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers, of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be A, B, or C, as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. Marking of shipments shall be in accordance with FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Shape of chisel required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, shapes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the shapes which are normally procured to meet Federal Government requirements.

6.3 Note. Because of the extensiveness of changes to this document, the margins have not been marked with asterisks to indicate where changes from the previous issue have been made.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969-393-067/S-2,32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document-(s) or serve to amend contractual requirements .

SPECIFICATION

GG-C-00289a(DSA-DM) FSN

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.