FED. SUP CLASS
6640



PART NUMBER	CAPACITY *µL (λ) (±%)
MS35986-5	1 ± 1.0
MS35986-6	5 ± 0.5
MS35986-1	10 ± 0.5
MS35986-2	25 ± 0.5
MS35986-7	

CANCELLED
CANCELLED AFTER 9 OCT 1991
NO SUPERSEDING STANDARD

MATER ... ation DD-G-541,

DESIGN: ... capered top (2.5 ± 0.5 mm ... a safety bulb near the top end of ... diameter at the safety bulb shall be no more than 1 mm ... than the maximum outside of the tubing.

Outside edge of each end shall be slightly ground and fire polished.

Micropipets shall be calibrated with mercury "to contain" at 20°C.

Marking of capacity in lambdas and calibrations shall be permanently etched or fused on the pipet.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

(C) CANCELLATION

APPROVED 23 Dec 1959 REVISED (A) 14 July 1969 (B) 21 OCT 1983 (C) 9 OCT 1991

P.A. EA; Other Cust MS 82	INTERNATIONAL INTEREST	TITLE: PIPET, VOLUMETRIC, MICRO Kirk design	MILITARY STANDARD MS35986
Procurement Specifications: ASTM E193	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE. CANCELLATION

Project Number 6640-1728-04

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

REVIEWER ACTIVITIES: ARMY - ME, NAVY - YD

USER ACTIVITIES: ARMY - ME

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.