
NOTICE.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS
6640



PART NUMBER	APPROXIMATE DIMENSIONS millimeters		
	LENGTH	WIDTH	THICKNESS
MS36317-1	114	114	3.2

MATERIAL: Glass

DESIGN: The ruled area shall be a 100 millimeter square subdivided into 10 millimeter squares, centered on the plate. All diagonal and the center four squares shall be subdivided into 9 equal squares each.

Lines may be engraved, etched or photo etched.

NOTES:
1. The counting plate is used with the Quebec Colony Counter MS36316-1.
2. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED, 8 SEP 1966 REVISED

P.A.	MU	TITLE	MILITARY STANDARD
Other Cust	BuMed MAA	PLATE, COUNTING, BACTERIOLOGICAL	MS36317
Procurement Specification GG-C-611		SUPERSEDES:	SHEET 1 OF 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:38Z
Check the source to verify that this is the current version before use.