FED. SUP CLASS
6630

PART NUMBER	STANDARD BUFFER SOLUTION pH VALUE
MS36453-1	1.00 ± .02
MS36453-2	2.00 ± .02
MS36453-3	3.00 ± .02
MS36453-4	4.01 ± .01
MS36453-5	5.00 ± .02
MS36453-6	6.00 ± .02
MS36453-7	6.840 ± .005
MS36453-8	7.00 ± .01
MS36453-9	7.384 ± .005
MS36453-10	8.00 ± .02
MS36453-11	9.00 ± .02
MS36453-12	10.00 ± .01
MS36453-13	11.00 ± .01

MATERIAL: Buffer solutions shall be prepared from ACS grade salts.

Bottles shall be of polyethylene.

PURPOSE: To provide a precise means for standardizing pH meters to a known reference.

COMPONENTS: Parts (Sets) shall consist of six polyethylene pint bottles of pH solution and a plastic dispenser.

Replacement pint bottles of pH solution are identified by the addition of suffix "A" after the part number, such as MS36453-1A for one pint of 1.00 pH solution.

NOTES:
1. Buffer solutions shall be supplied in one pint bottles and shall be standardized at 25°C.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.

APPROVED 7 OCTOBER 68 REVISED (A) 28 JAN 1982

USER ACTIVITIES: ARMY - CE
NAVY - NC

REVIEWER ACTIVITIES: ARMY - MD

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(A) ENTIRE STANDARD REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		STANDARD SOLUTION SET, HYDROGEN ION METER, PRE-MIXED	MS36453
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6630-0191
REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23117:412
Check the source to verify that this is the current version before use.