METRIC

A-A-50802A
22 June 1992
SUPERSEDING
A-A-50802
17 April 1981

COMMERCIAL ITEM DESCRIPTION

COUNTER, BLOOD CELLS, DIFFERENTIAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an eight unit differential blood cell counter with totalizer, suitable for use in routine white blood cell counts.

Salient characteristics:

The counter shall be a manually operated, automatic combination adding machine and percentage calculator consisting of eight counting units, each independently operated by a key and joined together in a line. The counter shall have individual amount windows for each key, a totalizer, a device which audibly signals the totalizer counting of each 100 key depressions, a removable blood cell chart, and a reset device.

The individual amount windows shall be oriented in the counter housing such as to be read when placed on a table while an operator is sitting or standing. As viewed by the operator, the windows shall register to 9, 9, 9, 99, 999, 99 and 99 respectively. As an alternate, each window may register to 999. The number of times each individual key is depressed shall register independently in its amount window. A single complete stroke on each key shall result in an accurate count without any evidence of double counting, skipping or coasting. The totals of the individual keys shall automatically record percentages when a total of 100 cells or multiples of 100 cells have been counted.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

The totalizer shall register the total number of times all of the eight individual keys have been depressed. The counter shall indicate when the totalizer has reached a count of 100 or multiples of 100 by a signal from a bell or other audible device. One revolution or operation of the reset device shall simultaneously reset all of the key counters and the totalizer to zero.

The counter body and gears shall be constructed of metal, and all parts subject to extreme wear shall be case hardened.

The counter shall be designed for recording differential white blood cell counts. The counter shall be provided with a removable classification strip or blood cell chart. The blood cell chart shall be made of a durable material and shall show the name and illustration of the particular blood cell type to be counted on each key. Each blood cell type shall be permanently and legibly illustrated in color, simulating morphological characteristics of the blood cells. The chart shall be fixed to the body of the counter in a readable position for the operator. Each blood type illustration shall be aligned on the counter with the specific window and key combination to be used for the counting.

The counter mechanism shall be enclosed in a sturdy metal housing, and shall be mounted on a base that prevents the counter from creeping or tipping during usage.

The overall dimensions of the counter shall not be greater than 30L X 12-1/2 x 12-1/2 cm. The weight of the counter shall not exceed 3 Kg.

The counter shall be supplied with detail instructions for operation and maintenance.

Workmanship. The blood cell counter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One blood cell counter, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the counter in accordance with commercial practice. Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Labeling. Labeling shall meet the requirements of the Food and Drug Administration and the Food, Drug and Cosmetic Act. The labels of the immediate container and the unit package of supply shall be those usually furnished by the supplier as in commercial practices. Addition to those labels shall be limited to the application of the NSN so the unit package, is supplied, and the expiration date to the immediate container (ship labeling is acceptable). When the unit package is the shipping container, additional labeling may be applied.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6640-00-418-8010	COUNTER, BLOOD CELLS, DIFFERENTIAL

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1535

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 5, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-50802A	22 JUNE 1992

3 DOCUMENT TITLE

COUNTER, BLOOD CELLS, DIFFERENTIAL

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a NAME *(Last, First, Middle Initial)*	b. ORGANIZATION	
c ADDRESS *(Include Zip Code)*	d TELEPHONE *(Include Area Code)* (1) Commercial (2) AUTOVON *(If applicable)*	7. DATE SUBMITTED *(YYMMDD)*

8 PREPARING ACTIVITY

a NAME	b TELEPHONE *(Include Area Code)* (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: MSE	(215)737-2870	444-2870
c ADDRESS *(Include Zip Code)* 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete* 98-270

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.