A-A-50807
April 17, 1981

COMMERCIAL ITEM DESCRIPTION

CYLINDER, GRADUATED, LABORATORY, GLASS

This commercial item description shall be used in preference to Federal Specification NNN-C-940.

This commercial item description covers requirements for glass graduated cylinders, suitable for laboratory purposes.

Salient characteristics:

Graduated cylinders shall be in the following styles: Style I (with pour-out lip), Style II (with standard-taper, stopper) and Style III (with pour-out lip and reinforced top).

Shall be cylindrical, graduated cylinders with hexagonal supporting base, having specified capacities and suitable for use in pharmacies and laboratories.

Shall be fabricated of borosilicate glass conforming to ASTM Standard E-438, Standard Specification for Glasses in Laboratory Apparatus.

Shall be permanently and legibly calibrated in accordance with NBS Circular 602, Testing of Glass Volumetric Apparatus.

Main graduations shall be suitably inscribed in Arabic Numerals.

Shall be calibrated and marked as follows:

To Deliver (TD) @ 20°C. (for Style I)
To Contain (TC) @ 20°C. (for Styles II & III)

Workmanship. Workmanship shall be first class throughout. Cylinders shall be free from defects which detract from their appearance or impair their serviceability.

Packaging and packing quantities. Unit. Each (EA). Each cylinder as specified, constitutes one unit. Intermediate package and shipping containers shall contain the quantity as furnished commercially.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Reference standard. Where standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of the description and conforms to the producers' own drawings, specifications, standards and quality assurance practices and is the same product offered for sale in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements. The offeror/contractor shall comply, and be responsible for compliance by their subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder, and the requirements of all other applicable Federal and State statutes and regulations.

Recovered materials. Offerors/contractors are encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging and packing. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof, under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from the receiving activity, and conforms to applicable carrier's rules and regulations.

NOTE: The following National Stock Numbers are representative of the items covered by this document: 6640-00-419-7000 and 6640-00-419-9900.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:
DoD - MB

Agent:
DLA - DM

Project No. 6640-1102

95831
95832

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.