A-A-50808
August 5, 1981

COMMERCIAL ITEM DESCRIPTION

URINOMETER, SQUIBB AND MICRO

This commercial item description shall be used in preference to Federal Specification GG-U-681b.

This commercial item description covers a squibb and a micro urinometer for use in determining the specific gravity of urine.

Salient characteristics:

The urinometers shall consist of a graduated glass hydrometer and an ungraduated glass cylinder. The cylinder shall have a depth approximating the length of the hydrometer and a diameter of at least 1/3 greater with a circular or hexagonal base of sufficient diameter to prevent tipping over. The hydrometer shall consist of a cylindrical tube or stem, a ballast bulb and a buoyancy bulb located immediately above the ballast bulb. The hydrometer shall float freely within the cylinder. A graduated scale shall be secured within the upper stem of the hydrometer and shall be inscribed on white paper with nonfading ink. All graduations shall be clear and sufficiently distinct to be read without the aid of a magnifying glass. The squibb urinometer shall have a scale range of 1.000 to 1.060 divided into six major divisions by graduating lines numbered with arabic numerals. The micro shall provide readings with approximately 16 ml of urine. The micro urinometer shall have a scale range of 1.000 to 1.040 divided into five major divisions by graduating lines numbered with arabic numerals. For both types, the major graduations shall be subdivided into halves and tenths. The scale shall be calibrated at 60°F (15.56°C) and shall be so labeled.

Workmanship. Workmanship shall be first class throughout. The urinometers shall be free from defects which detract from their appearance or impair their serviceability.

Packaging and packing quantities. Unit. Each (EA). Each urinometer, as specified, constitutes one unit. Shipping container shall contain ten units.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of the description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product offered for sale in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements. The offeror/contractor shall comply, and be responsible for compliance by their subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder, and the requirements of all other applicable Federal and State statutes and regulations.

Recovered materials. Offerors/contractors are encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, and packing. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and conforms to applicable carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the items described herein:

Squibb Urinometer:	
Mercer Glass Works Inc.	Part No. 2430S or 2430M
Clay Adams, Division of Becton Dickinson Co.	Part No. 1520
LaPine Scientific Company	Part No. D03176
Micro Urinometer:	
Mercer Glass Works Inc.	Part No. 2420SS
Clay Adams, Division of Becton Dickinson Co.	Part No. 1525
LaPine Scientific Company	Part No. D03181

NOTE: The following National Stock Numbers are representative of the items covered by this document: 6630-00-447-0000, 6630-00-299-9832

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:
DoD - MB

Agent:
DLA - DM

95212

Project No. 6630-0335

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.