A-A-50809
August 5, 1981

COMMERCIAL ITEM DESCRIPTION

SHAKING MACHINE, LABORATORY, SLIDE, 115 Volt, AC-DC

This commercial item description shall be used in preference to Military Specifications MIL-S-3678 and MIL-S-003678A (DSA-DM).

This commercial item description covers a laboratory shaking machine for simulating hand shaking of micro slides, antigen flasks and other equipment for flocculation tests in connection with cardiolipin syphilis tests.

Salient characteristics:

The shaking machine shall consist of a suitably finished steel base, a housing, a continuous duty, 115 Volt, AC-DC, universal motor with a worm gear and a rubber covered steel or aluminum floating platform approximately 13 inches square.

The shaking machine shall have overall dimensions of 13 inches square by 7-1/2 inches high and shall provide variable speeds from a low of 52 ± 31 RPM to a nominal high of 180 RPM. An automatic cutoff timing device shall be provided with a dial graduated to indicate 0 to 28 minutes in 1/2 minute subdivisions. At least every 5 minute interval shall be numbered. The shaking machine shall be provided with a three-wire electric power cord with a three-prong male connector, meeting the requirements of ANSI C73.11.

Workmanship. Workmanship shall be first class throughout. The shaking machine shall be free from defects which detract from its appearance or impair its serviceability.

Packaging and packing quantities. Unit. Each (EA). Each shaking machine, as specified, constitutes one unit. Shipping container shall contain one unit.

Reference standard. Where a standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of the description and conforms to the producers' own drawings, specifications, standards and quality assurance practices, and is the same as the product offered for sale in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements. The offeror/contractor shall comply, and be responsible for compliance by their subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder, and the requirements of all other applicable Federal and State statutes and regulations.

Recovered materials. Offeror/contractors are encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, and packing. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity plus reshipment from the receiving activity, and conforms to applicable carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

For Information only. The following product, as of the date of this document, are purported by the following companies to meet the salient characteristics of the item described herein:

Clay Adams, Division of Becton Dickinson and Company	Part No. 1213
LaPine Scientific Company	Part No. 30-32

NOTE: The following National Stock Number is covered by this document:

6640-00-438-9600

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:
DoD - MB

Agent:
DLA - DM

55385

Project No. 6640-1110

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.