A-A-50814A
31 January 1989
SUPERSEDING
A-A-50814
August 5, 1981

COMMERCIAL ITEM DESCRIPTION

BASIN, WASH

Corrosion-Resisting Steel

The General Services Administration has authorized the use of this Commercial Item Description in preference as a replacement for Federal Specification RR-B-120 which is cancelled.

This Commercial Item Description covers a wash basin suitable for use in hospitals and first aid stations.

Salient characteristics:

The wash basin shall be a conventional, deep, round, clinical solution basin, fabricated of corrosion-resisting steel of the American Iron and Steel Institute (AISI) 300 series, suitable for use in hospitals and first aid stations. The basin shall be furnished in the following sizes:

Size I - 7 ± 1/2 quarts (6.6 ± .55 liters)

Size II - 9 ± 1/2 quarts (8.5 ± .55 liters)

The basin shall be formed seamlessly from one piece of 0.031 inch thick (No. 22, U.S. Standard) steel, which shall not be reduced below 0.020 inch thick at any point in the finished basin. The entire top of the basin shall be flanged outward ending in a circular bead underneath.

The Size I basin shall have an overall diameter of 13-5/8 ± 1/8 inches, a diameter of 11-13/16 ± 1/8 inches when measured 3/8 inches below the top and a diameter at the bottom measuring 9-1/2 ± 1/8 inches. As an alternate, the diameter when measured 3/16 inches below the top shall be 12-1/16 ± 1/8 inches and a diameter at the bottom of 9-27/32 ± 1/8 inches.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

The Size II basin shall have an overall diameter of 14-1/2 ± 1/8 inches, a diameter of 12-9/16 ± 1/8 inches when measured 3/8 inch below the top and a diameter at the bottom measuring 10-1/2 ± 1/8 inches.

The basin shall be stackable for storage and shall rest evenly on a flat surface, without rock or wobble, while in use.

The top edge of the basin shall be free of sharpness or burrs. The basin surface shall be free of pits, scales, scratches, or grinding and trimming marks. The basin shall have a commercial No. 4 finish, or better. It shall be free of dirt, grease, graphite, metallic particles, or other foreign particles.

Workmanship. The wash basin shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One basin, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one wash basin as specified in this description.

Defects.

Basin is not fabricated seamlessly from American Iron and Steel Institute (AISI) Series 300 corrosion-resisting steel.

Basin top is not flanged outward with a circular bead underneath.

Basin does not stack properly.

Basin does not rest evenly on a flat surface without rock or wobble.

Top edge of basin is not free of sharpness or burrs.

Basin does not have a commercial No. 4 finish, or better.

Basin surface is not free of pits, scales, scratches, or grinding and trimming marks.

Basin is not free of dirt, grease, graphite, metallic particles, or other foreign particles.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one wash basin.

Tests

Capacity. With the basin sitting on a flat surface, it shall be filled with water from a graduated source. Having a capacity greater or less than specified, shall constitute a defect.

Leakage. With the basin sitting on a flat surface, it shall be filled with water and allowed to stand for a period of six hours. Any evidence of leakage shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-00-771-0105 (Size I)
6530-00-771-0225 (Size II)

NOTE: Size I basin shall be of a dimensional size to fit the basin holder of NSN 6530-00-708-1000, STAND, BASIN.

Size II basin shall be of a dimensional size to fit the holder of NSN 6530-00-708-1025, STAND, BASIN.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2022

Location: CID/BASIN/WASH/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions — Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50814A	BASIN, WASH , Corrosion-Resisting Steel

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* — Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* — Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* — Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 92 MAR

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTD
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.