A-A-50815A
24 March 1989
SUPERSEDING
A-A-50815
13 August 1981

COMMERCIAL ITEM DESCRIPTION

LIGHT, DESK

High Intensity

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-L-37560 which is cancelled.

This Commercial Item Description covers a small, high-intensity desk light suitable for use in hospital wards.

Salient characteristics:

The desk light shall be a miniature, glare-free, dual-intensity, high intensity type light suitable for use on a hospital ward desk. The light shall consist of a base and a lamphead, connected by an adjustable, variable height, articulating arm, and a power cord and plug.

The base dimensions shall not exceed 3 inches wide, 5 inches long and 2-1/2 inches high. The base shall contain a transformer and a three-position switch with "off", "low intensity," and "high intensity" positions. The base shall rest stably on a flat surface and shall have a slotted hole in the bottom for wall mounting. The light shall have two arms connected together which shall permit a wide range of vertical and horizontal adjustment. The minimum length of the lower and upper arms shall be 6 inches and 8 inches, respectively. The arms shall be fabricated from plated steel or corrosion-resisting steel tubing, either round or rectangular.

The lamphead shall consist of the reflector, the socket, and the lamp. The reflector shall have a diameter of 4 to 4-1/2 inches. The lamp shall be a commercial number 1133 incandescent lamp. The light shall be supplied with a lamp and two spares.

AMSC N/A FSC 6230

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

The power cord shall be a minimum of 5 feet long, measured from the base of the light, and shall be terminated in a three-prong plug. If the unit is double insulated, the power cord may be terminated in a two-prong plug. The cord shall enter the base through a strain-relief bushing. The cord and plug shall conform to Underwriters' Laboratories, Inc. (UL) 153. All other wiring and electrical components shall meet the applicable requirement of the UL. The light shall be tested for dielectic withstand and current leakage per UL 153 and UL 1262, respectively. Proof of compliance with UL requirements shall consist of UL listing of the item or a report from a nationally recognized testing laboratory showing compliance with the requirements of UL.

The light base shall be covered by a colored vinyl laminate or a brass, nickel, chromium, or cadmium plating. The interior of the reflector shall be painted white; the exterior shall be finished to match the base. All other parts of the light shall be either corrosion resisting, or shall be appropriately finished to resist corrosion.

Adequate strain relief shall be provided throughout the light assembly, effective in all positions of arm adjustment. Wire raceways shall be free of burrs and sharpness. All connections shall be firmly and securely made. Insulation shall be sound, with only essential stripping made for terminal connections.

Exposed metal surfaces of light shall not be bent, dented, or cracked. Exposed light surfaces shall not have burrs or sharpness, or be scratched or marred.

Workmanship. The desk light shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). Each light, with a lamp and two spares, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one desk light as specified in this description.

DEFECTS

Any component does not function as specified.

Extension arms are not easily adjustable.

Lamphead is not easily adjustable.

Light does not remain fixed when set in any position.

Strain relief bushing does not hold cord securely in place.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

DEFECTS (continued).

Light base does not have slotted hole for wall mounting.

Exposed metal surface is not free from bends, dents, or cracks.

Exposed light surface is not free from burrs or sharpness.

Internal wiring is bare of insulation.

Wire raceway is not free from burrs or sharp edges.

Adequate strain relief is not provided throughout light.

Light finish is not as specified.

Light surface is scratched or marred.

Light cord is less than 5 feet in length.

Two spare lamps are not furnished.

Sampling for tests. Sampling for tests (including dimensional tests), shall be conducted in accordance with MIL-STD-105, with an AQL of 0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one desk light.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests

Dielectric withstand. The light shall be tested for dielectric withstand as specified in UL 153 (Portable Electric Lamps). Any signs of dielectric breakdown shall constitute a defect.

Leakage current. The light shall be tested for leakage current as specified in UL 1262 (Laboratory Equipment). Leakage current greater than 0.5 milliamperes shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6230-00-809-4284

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6230-1010

Location:
CID/LIGHT/DESK/H-13-33

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50815A	LIGHT, DESK, High Intensity

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.