A-A-50816A
30 May 1989
SUPERSEDING
A-A-50816
13 August 1981

COMMERCIAL ITEM DESCRIPTION

JAR, SURGICAL DRESSING

Corrosion-Resisting Steel

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Interim Federal Specification RR-J-00131b and Federal Specification RR-J-131a which are cancelled.

This Commercial Item Description covers a corrosion-resisting steel surgical dressing jar and cover designed for use in medical treatment facilities.

Salient characteristics:

The surgical dressing jar shall have a capacity of 4-1/2 quarts plus or minus 1/2 quart. The diameter of the jar shall be 6-1/2 inches plus or minus 1/2 inch. The design of the jar shall facilitate easy cleaning. The jar shall sit evenly on a flat surface without rock or wobble.

The jar and cover shall each be formed seamlessly from one single piece of corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series, with a thickness of not less than 0.025 inch (No. 24, United States Standard Gage), which shall not be reduced below 0.018 inch in thickness at any point in the finished jar or cover.

The cover shall fit snugly, but not tightly, over the body, so that the cover may be removed with one hand without the need for holding the jar in the other hand. The cover shall have an AISI series 300 corrosion-resisting steel handle, either square or rounded. Finger clearance shall be adequate. The handle shall be securely welded or spot welded to the cover top. Welding shall be thorough, sound, and free from pits, flashes, and discoloration.

The jar, cover, and handle shall be free of sharp edges, dents, cracks, tears, ripples, or deformities. Jar and cover shall have a commercial No. 4 finish or better.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

The jar, cover, and handle shall be free of chips, oil, grease, dirt, or other foreign matter.

The conditions for production and/or marketing of this item, if required by the Food and Drug Administration, shall be specified in the procurement of this item.

Workmanship. The surgical dressing jar shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One jar, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical dressing jar, as specified in this description.

Defects.

Surgical dressing jar is not complete with all component parts.

Jar, cover, or handle is not fabricated from AISI series 300 corrosion-resisting steel.

Jar, cover, or handle is not free of sharp edges, dents, cracks, tears, ripples, or deformities.

Jar does not sit evenly, without rock or wobble, on a flat surface.

Jar or cover is not free of course grind marks, scratches, or mars.

Jar or cover finish is not a No. 4 Commercial finish or better.

Jar, cover, or handle is not free of chips, oil, grease, dirt, or other foreign matter.

Sampling for tests. Sampling for tests including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one surgical dressing jar.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Tests

Capacity Test. A surgical dressing jar sample shall be filled with water to 1/4 inch of the top from a known measured source. A capacity less than 4 quarts or greater than 5 quarts shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-782-7075

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2041

Location: CID/JAR/SURGICAL/H13-33

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide sug- -tions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and m In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, wa too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN. DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-50816A	JAR, SURGICAL DRESSING, Corrosion-Resisting Steel

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.