A-A-50818
October 26, 1981

COMMERCIAL ITEM DESCRIPTION

TIP, PIPET, PLASTIC, 0.2 ml, Disposable, 1000s

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable 200 microliter capacity pipet tip for use with the Medical Laboratory Automation, Inc. (MLA) system of pipettors.

Salient characteristics:

Shall be a 200 microliter capacity pipet tip that shall be designed to fit the nozzles of MLA Precision pipets having volumetric sizes of 10 to 200 microliters.

Shall be fabricated of polypropylene having non-wetting characteristics for complete dispensing of a liquid sample, and having sufficient optical clarity for checking the level of a liquid sample that is drawn into the bore of the pipet tip.

The dimensions of the pipet tip shall be such that, in a vertical position, the level of liquid sample drawn into the bore (by the plunger action of an attached MLA Precision pipet) does not come in contact with the nozzle of the pipet.

Workmanship. The pipet tip shall be free from defects which detract from its appearance or impair its serviceability.

Packaging and packing quantities. Unit. Package (PG). One package containing one thousand tips, as specified, constitutes one unit. Shipping container shall contain ten units.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of the description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements. The offeror/contractor shall comply, and be responsible for compliance by their subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder, and the requirements of all other applicable Federal and State statutes and regulations.

Recovered materials. Offerors/contractors are encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, and packing. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and conforms to applicable carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:47Z
Check the source to verify that this is the current version before use.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the item described herein:

Medical Laboratory Automation, Inc. (MLA)	Catalog No. 4025
Labcon Inc.	Catalog No. LW 1060
Centaur Sciences	Catalog No. IPM 10-200
American Scientific Products	Catalog No. P5059-600
J & H Berge	Catalog No. X373447G
Cole Parmer Instrument	Catalog No. 6242-10-FB
Whittaker General Science	Catalog No. 1060-500

NOTE: The following National Stock Number is covered by this document:

6640-01-095-6886

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

56756

Project No. 6640-1149

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:47Z
Check the source to verify that this is the current version before use.