A-A-50825
December 2, 1981

COMMERCIAL ITEM DESCRIPTION

CHAMBER, COUNTING, BLOOD CELLS, HEMACYTOMETER

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-H-216C.

This commercial item description covers a counting chamber, suitable for use in counting red cells, white cells and platelets.

Salient characteristics.

The counting chamber shall be constructed of one-piece glass, with an open-type cell chamber and an H-shaped moat forming two counting areas with improved Neubauer rulings. The chamber shall be in accordance with National Bureau of Standards (NBS) Scientific Paper Number 507.

The ruling shall consist of lines ruled through a semi-transparent, bright metal-coated surface, which has been deposited on the glass counting area.

Workmanship. The counting chamber shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One chamber, as specified, constitutes one unit.

Reference standard. Where standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the item described herein:

American Optical Corporation	Part No. 1490
C. A. Hausser and Son	Part No. 100

NOTE: The following National Stock Number is covered by this document: 6630-00-427-6850.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

53303

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6630-0333

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.