A-A-50832
March 8, 1982

COMMERCIAL ITEM DESCRIPTION

TUBE, FILTERING, Disposable

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable, 4 inch long filtering device for use in removing fibrin from centrifuged blood serum contained in a standard 13 millimeter (mm.) diameter test tube.

Salient characteristics:

Shall be a filtering device that is designed to fit a standard 13 mm. diameter test tube. As a minimum, shall have a nominal length of four (4) inches.

Shall consist of a clear, hollow, plastic tube (cylinder) with a porous filter that is fitted into its distal end.

Shall remove fibrin from centrifuged blood serum when inserted into a standard 13 mm. diameter test tube containing the blood serum. During its insertion and filtering action in the test tube, the device shall not be susceptible to plugging by fibrin material. In addition, the device shall operate with minimum difficulty so as to prevent possible contaminated serum from squirting out of the test tube.

The filtered blood serum shall be able to be poured out of the test tube through the plastic tube in the filtering device.

The filtering device shall not alter nor react with the chemical composition of the blood serum.

Workmanship. The filtering device shall be free from defects which detracts from its appearance or impairs its serviceability.

Unit. Package (PG). One package containing one hundred filters, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the items described herein:

Acculab	Cat. No. 51-0063 Accu-Filter
American Scientific Products	Cat. No. P5191-4 Filter Sampler
Curtin Matheson Scientific	Cat. No. 061-549 Filter Sampler
Glasrock	Cat. No. FS413 Filter Sampler
VWR Scientific	Cat. No. 28295-603 Filter Sampler
Whittaker General Scientific	Cat. No. 113736 Accu-Filter

NOTE: The following National Stock Number is covered by this document: 6640-00-262-0999

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

PGC 56764

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6640-1166

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.