
FED. SUP CLASS
6640



MS35979-3 (Thoma) White Cell | MS35979-4 (Thoma) Red Cell | MS35979-5 (Sahli)

PART NUMBER	NOMENCLATURE
MS35979-1	CANCELLED W/O REPLACEMENT
MS35979-2	CANCELLED W/O REPLACEMENT
MS35979-3	PIPET, BLOOD DILUTING
MS35979-4	PIPET, BLOOD DILUTING
MS35979-5	PIPET, BLOOD DILUTING
COMPONENT PART	
MS35979-6	MOUTHPIECE, PIPETING

PART NUMBER	NUMBERED VOLUME INTERVALS FOR DILUTING
MS35979-3	0.5 (1-20 dilutions) 1.0 (1-10 dilutions) 11.0 (total capacity)
MS35979-4	0.5 (1-2000 dilutions) 1.0 (1-1000 dilutions) 101.0 (total capacity)
Ⓑ MS35979-5	20 mm^3 (0.02 mL)(total capacity)

MATERIAL: MS35979-3, MS35979-4, and MS35979-5 shall be of non-defective and bubble free glass in accordance with N.B.S. Circular 602.

MS35979-6 shall be of plastic or bone.

Ⓑ DESIGN: MS35979-3, white cell diluting pipet and MS35979-4, red cell diluting pipet shall be cylindrically shaped with a uniform bore and a blown bulb of proper size, containing a glass mixing bead located between the capillary tube and the last engraved mark.

MS35979-5 shall be cylindrically shaped with a uniform bore and gradual tapered delivery tip which shall be smoothly ground or fire polished. Pipet shall have a fused-on single graduation mark and shall be calibrated to contain 20 cubic millimeters (0.02 mL).

MS35979-6 shall be wedge-shape, designed to fit rubber tubing which shall be attached to the above blood diluting pipets.

Ⓑ DENOTES CHANGES

P.A. EA Other Cust MB MS 99	INTERNATIONAL INTEREST	TITLE PIPET, BLOOD DILUTING Thoma and Sahli designs	MILITARY STANDARD MS35979
Procurement Specifications See note 2	SUPERSEDES:		SHEET 1 OF 2

PROJECT NUMBER 6640-1407-2

DD FORM 1 MAY 73 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

APPROVED 23 DEC 1959 REVISED Ⓐ 14 JUL 1969 Ⓑ 16 MAR 1987

REVIEWER ACTIVITIES: ARMY - MD, ME
NAVY - OS, YD
AIR FORCE - 82

USER ACTIVITIES:

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.

NOTES:

1. Pipets shall not be furnished with rubber tubing and mouthpiece.

Ⓑ 2. Acquisition document for MS35979-3 and MS35979-4 is Commercial Item Description (CID), A-A-50812.

3. In the event of a conflict between the text of this standard and the references cited herein, the text of this standard shall take precedence.

4. Referenced Government documents of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DODISS) specified in the solicitation form a part of this standard to the extent specified herein.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED REVISED Ⓑ For changes see pages 1 and 2.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MB MS 99		PIPET, BLOOD DILUTING Thoma and Sahli designs	MS35979
Procurement Specifications See note 2	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1407-2

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.