

FED. SUP CLASS
6640



MS36595-1



MS36595-2



MS36595-3

PART NUMBER	NOMINAL OVERALL HOUSING DIMENSIONS (MM)		
	LENGTH	WIDTH	HEIGHT
MS36595-1	158.7	127	152
MS36595-2	158.7	89	165
MS36595-3	290	310	240

MATERIAL: MS36595-1 housing shall be of heavy cast iron or die cast aluminum with a baked on enamel finish.

MS36595-1 oscillating cup and support string shall be of neoprene.

MS36595-2 housing shall be enameled steel case with an aluminum platform.

MS36595-2 pipet holder shall be of corrosion-resisting spring steel or nickel plated bronze spring clips.

MS36595-3 housing shall be made of steel in accordance with ASTM A505 with enamel or lacquer finish. Four rubber feet shall be provided. The top may be steel or aluminum.

DESIGN: MS36595-1 shall be of the reciprocating cup type with a built-in speed rheostat and pressure actuated switch. It shall be designed to operate with instant touch starting as container is placed in oscillating cup. It shall accommodate micro size to 40 mm test tubes, all sizes of centrifuge tubes, and volumetric flasks up to 100 mL.

MS36595-2 shall consist of a vertically balanced electric synchronized vibrator enclosed in a compact housing with an amplitude adjustable knob providing a shaking pattern independent of voltage fluctuations. It shall accommodate a maximum of two pipets.

MS36595-3 shall consist of a 1/30 hp split phase induction motor with V-belt transmission having a variable pitch pulley system, speed control, locking knob and on/off switch in front panel. The platform shall be 36 x 31 cm with 168 holes for mounting flask holders.

(A) ENTIRE STANDARD REVISED

APPROVED 15 April 1970 REVISED (A) 8 NOV 1983

USER ACTIVITIES: ARMY - DM, MD, MI
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		SHAKING MACHINE, LABORATORY, RECIPROCATING	MS36595
Other Cust: MS 99			
Procurement Specifications: NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0779

Source: https://assist.dla.mil -- Downloaded: 2026-04-23T17:47Z
Check the source to verify that this is the current version before use.

FED. SUP CLASS
6640

Rubber suction feet or sponge strips shall be attached to the base of the shakers to provide stability.

Shakers shall be supplied with a 3 wire cord and plug for 115V, 50/60 cycles AC.

When specified, the shaking machines shall conform to the electro-magnetic characteristic requirements and test limits for Class IIB equipment as specified in MIL-STD-461.

All electrical components shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.
3. FSCM: 13134, 15481, 59728, 64484, and 91290.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

APPROVED REVISED 8 NOV 1983

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		SHAKING MACHINE, LABORATORY, RECIPROCATING	MS 36595
Procurement Specifications NONE	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0779

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:44Z
Check the source to verify that this is the current version before use.