A-A-54266
12 October 1990

COMMERCIAL ITEM DESCRIPTION

SURFACER, OPHTHALMIC LENS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a table top mounted semi-automatic ophthalmic lens surfacer designed and constructed for polishing and fining spherical and toric, glass and plastic spectacle lenses.

Salient Characteristics:

Shall be a table top semi-automatic ophthalmic lens surfacing machine.

The unit shall surface and polish glass and plastic lenses.

The contact pressure between the lap and lens shall be continuously adjustable and applied through spring loaded pressure rods with the conversion capability to operate using line air pressure. A contact pressure indication gage shall be installed in plain view to the operator.

Shall be a twin spindle, dual speed machine which operates at nominal speeds of 570 and 875 inches per minute.

The unit shall utilize a self contained pump and epoxy coated tank with a nominal fluid capacity of 32 ounces.

The tube assembly shall be equipped with Aircraft type, sealed, self-aligning bearings.

All bearings, bushings, interior components and assemblies shall be completely protected and sealed from the polishing and fining compound.

Shall use "push-button" timer with automatic reset that eliminates the need for continual setting. An elapsed time meter to provide automatic shut-off and total running time shall be included.

The unit shall provide for easy access to all moving parts.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:31Z
Check the source to verify that this is the current version before use.

The maximum allowable weight of the entire unit shall be 350 lbs.

Power requirements.

The surfacer shall operate using a single phase motor not less than 1/6 H.P., 3.2 amp motor using a nominal 110 volt, 60 hertz power supply.

Workmanship. The surfacer shall be free from defects which detract from its appearance or impair its serviceability. Metal parts shall be free from burrs, pits and deformations. Finished surfaces shall be free of chips, scratches and abrasions.

Unit. Each (EA). One surfacer, as specified, constitutes one unit.

Commercial Manuals.

The contractor shall provide two (2) copies each of the installation instructions, user/operator manuals, user maintenance manuals, service manuals, parts price lists and wiring schematics/diagrams.

System Warranty.

The contractor shall provide for a standard commercial warranty for the surfacer. The warranty shall provide for unlimited technical support, on-site service, repair and parts for a period of at least one (1) year.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the currentversion before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with be Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6540-01-227-3628 SURFACER, OPHTHALMIC LENS

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6540-1251

Location: ENABLE/SURFACEC/S3C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.