A-A-54270
28 February 1992

COMMERCIAL ITEM DESCRIPTION

ELEVATORS, DENTAL, SURGICAL, GENERAL REQUIREMENTS FOR

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-E-1161 which is cancelled.

This commercial item description covers dental elevators for use in oral surgical procedures.

Salient Characteristics:

Commercial item description sheet. The surgical elevators shall be in accordance with the requirements specified herein and the applicable commercial item description sheet (slash sheet-see Notes).

Material.

Operative ends. The operative end and shank shall be fabricated of class 4, type 440A stainless steel in accordance with ASTM Designation F899 for composition only.

Handles. Unless otherwise specified in the commercial item description sheet for the applicable elevators, the handles shall be fabricated of plated copper alloy C26000 or plated copper alloy C33000 in accordance with ASTM Designation B135 for composition only. One piece instruments shall be fabricated of the same material as the operative ends as specified herein.

Hardness. The operative ends shall have a Rockwell hardness of C49 to C56 when tested in accordance with ASTM Designation E18 or E92 and E140. The hardness of the shanks shall be suitable for the purpose intended.

Design and dimensions. The design and dimensions of the elevators shall be as specified in the detailed commercial item description sheet for the applicable elevator number or letter.

Construction. The elevator shall consist of an operative end (blade) and handle. The handle of two or three piece construction shall be permanently secured to the handle by

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

brazing or press fitted to the handle with no crevices evident, and cemented. The interface between the shank and the handle shall be smoothly blended with no crevices evident.

Finish.

Operative ends. The operative ends shall have a bright finish with a finely honed finish on the inside faces, except as otherwise specified in the commercial item description sheet.

Handles. Unless otherwise specified in the commercial item description sheet for the applicable elevator, the finish shall be satin or matte finish. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with a fine abrasive under pressure. The shank retaining portion of the handles other than the one-piece instruments shall have a bright finish.

Corrosion resistance.

Operative ends. The operative ends shall meet the copper sulfate and boil test requirements as specified in ASTM Designation F1089.

The entire instrument shall meet the requirements of the boil test requirements as specified in ASTM Designation F1089.

Identification marking. Each elevator shall be legibly and permanently marked with the applicable instrument number or letter. Elevators of foreign manufacture shall bear the name of the country of origin and the word stainless.

Individual package. Each elevator shall be wrapped in neutral tissue and securely packaged to prevent damage to the operative ends during shipment and handling.

Workmanship. The entire elevator shall be free from defects such as cracks, coarse grind marks, pits, feathers or burrs which detract from their appearance or impair their serviceability, proper function and intended use.

Unit. Each (EA). One elevator, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, test purposes shall be one elevator. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1. Acceptance number for tests shall be zero (0) for all samples.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements..

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

NOTES:

a. Contracting officer should select the preferred options permitted herein, and include the following information in procurement documents:

(1) Title, number and dated of the commercial item description.

(2) Title, number and date of the applicable commercial item description sheet as follows.

A-A-54270/1	ELEVATOR, ROOT, No. 34S, No. 92, No. 301, No. 304 and Stout A.
A-A-54270/2	ELEVATOR, ROOT, Cogswell A and Cogswell B.
A-A-54270/3	ELEVATOR, ROOT, Seldin No. 1L and No. 1R.
A-A-54270/4	ELEVATOR, ROOT, Miller No. 73 and No. 74.
A-A-54270/5	ELEVATOR, ROOT, West No. 9, No. 9L and No. 9R.
A-A-54270/6	ELEVATOR, ROOT, Heidbrink Ejector, No. 1, No. 2 and No. 3.
A-A-54270/7	ELEVATOR, ROOT, Winter, No. 122 and No. 123.
A-A-54270/8	ELEVATOR, PERIOSTEAL, Molt No. 9, Allen No. 9A and Seldin No. 23.
A-A-54270/9	ELEVATOR, MALAR, Ivy.

b. Copies of Commercial Item Description A-A-54270 and Commercial Item Description Sheets may be obtained by writing Commanding Officer, Navy Publications and Forms Center, (code 1032), 5801 Tabor Avenue, Philadelphia, PA 19120.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

A-A-54270

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2421
LOCATION: KD3/Elevator

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54270	2. DOCUMENT DATE (YYMMDD) 92-02-28

3. DOCUMENT TITLE
ELEVATORS, DENTAL, SURGICAL, GENERAL REQUIREMENTS FOR

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

A-A-54270/1
28 February 1992

COMMERCIAL ITEM DESCRIPTION SHEET

ELEVATOR, ROOT, Maxilla

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-E-1161 which is cancelled.

The complete requirements for procuring Elevator, Dental, Surgical shall consist of this document and Commercial Item Description A-A-54270.

Salient characteristics:

In addition to the salient characteristics of A-A-54270 the following shall apply.

Design and dimensions. This commercial item description sheet covers the following root elevators:

No. 34S
No. 92
No. 301
No. 304
Stout A

The design and dimensions of the root elevators shall be in accordance with the applicable drawing and purpose listed in the table herein.

ELEVATOR	DPSC DRAWING	PURPOSE
No. 34S	20347	Used to elevate, luxate teeth from the maxilla.
No. 92	24416	Used for luxation and removal of mandibular and maxillary roots and teeth.
No. 301	20699	Used primarily on the maxilla for the removal of root fractured at the gingival line.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

ELEVATOR	DPSC DRAWING	PURPOSE
No. 304	24515	Used to remove roots, root tips and spicula.
Stout A	2190	Used to remove teeth and roots by wedge action.

NOTES:

a. Copies of Commercial Item Description A-A-54270 and Commercial Item Description Sheets may be obtained by writing Commanding Officer, Navy Publications and Forms Center, (code 1032), 5801 Tabor Avenue, Philadelphia, PA 19120.

b. National Stock Number and Item Identification.

National Stock Number	Item Identification
6520-00-524-2550	ELEVATOR, ROOT, No. 34S
6520-01-057-7425	ELEVATOR, ROOT, No. 92
6520-00-524-3050	ELEVATOR, ROOT, No. 301
6520-01-104-0419	ELEVATOR, ROOT, No. 304
6520-00-524-6500	ELEVATOR, ROOT, Stout A

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2422
LOCATION: KD3/54270-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54270/1	2. DOCUMENT DATE (YYMMDD) 92-02-28

3. DOCUMENT TITLE
ELEVATOR, ROOT

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.