A-A-5427.
7 November 1990

COMMERCIAL ITEM DESCRIPTION

CANNULA, TRACHEOSTOMY, MODIFIED, JACKSON

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-36643A(1) which is cancelled

This commercial item description covers two types of tracheostomy cannulas of the modified Jackson design, with or without an endotracheal adapter suitable for insertion into the trachea

Salient characteristics

General Tracheostomy cannula shall be furnished in the size and type designated for each specified National Stock Number (See Note on page 5).

Sizes (O D. of Outer Tube)

2 - 0 236 inch
3 - 0 274 inch
4 - 0 314 inch
5 - 0.351 inch
6 - 0 393 inch
7 - 0 440 inch
8 - 0 474 inch

Types

I - Without tracheal tube adapter
II - With tracheal tube adapter

Material The inner and outer tube neck plate, swivel lock plate, inner tube anchor plate, swivel lock and tube adapter shall be Class 3 corrosion resistant steel in accordance with ASTM Designation F899 The stylet and bearing washer shall be fabricated from other suitable materials

Stylet (Obturator) The stylet shall be fabricated from half-hard tempered brass material, and shall be nickel or chromium plated

Bearing washer The washer shall be fabricated from non-toxic, durable, non-metallic material capable of repeated steam sterilization

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:31Z
Check the source to verify that this is the current version before use.

Hardness The cannula shall have a Rockwell Hardness B scale (HRB) of 85 to 95 when tested in accordance with ASTM Designation E18 or E92 whichever is applicable Conversion of hardness values shall in accordance with ASTM Designation E140

Style, design and dimension The cannula shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 24083

The stylet shall fit smoothly within the outer tube, shall blend through the distal opening of the outer tube without completely falling through, and shall not bind into the opening

Finish The finish of the cannula shall be mirror, satin or matte The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure

Passivation The cannula shall be properly passivated

Corrosion resistance For components fabricated from Class 3, no signs of corrosion shall appear when subjected to the Boil Test in accordance with ASTM Designation F1089 For nickel or chromium plated components, no signs of corrosion shall appear when salt spray tested in accordance with ASTM Designation B117

Identification marking and labeling Cannulas shall be legibly marked with the contractor's name or trademark and the word 'stainless'. Foreign manufactured cannulas shall bear the name of the country of origin.

Performance interchangeability The outer and inner cannula and stylet shall be interchangeable with all same size component parts, regardless of manufacturer, and shall have a proper fit with mating parts The same interchangeability and fit requirements shall also exist with the same size adapter

Workmanship The cannula shall be free from cracks, pits, burrs, sharp edges, scale rough areas, corrosion or any defects which detract from its appearance or impair its serviceability

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:31Z
Check the source to verify that this is the currentversion before use.

PRODUCT QUALITY LEVEL

Visual and dimensional examination Cannulas shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified

Unit Each (EA) One cannula of the type and size specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except, as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein Unit of product for examination purposes shall be one cannula Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination The cannulas shall be examined to determine compliance with specification requirements

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein The unit of product for test purposes shall be one cannula Sampling shall be Inspection level S-1 with an AQL of 1 0

Tests Tests shall be conducted to determined compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements.

Federal Food, Drug and Cosmetic Act Of the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

NOTE: The following National Stock Numbers, their respective Item Identifications, and Types are listed below

NSN	ITEM DESCRIPTION	TYPE
6515-00-518-5461	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 2	I
6515-00-518-5463	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 2, with 15 mm adapter	II
6515-00-518-5464	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 3	I
6515-00-518-5467	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 3, with 15 mm adapter	II
6515-00-518-5469	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 4	I
6515-00-518-5472	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 4, with 15 mm adapter	II
6515-00-518-5473	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 5	I

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:31Z
Check the source to verify that this is the current version before use.

6515-00-518-5474	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 5, with 15 mm adapter	II
6515-00-518-5475	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 6	I
6515-00-518-5477	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 6, with 15 mm adapter	II
6515-00-518-5478	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 7	I
6515-00-518-5481	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 7, with 15 mm adapter	II
6515-00-518-5494	CANNULA, TRACHEOSTOMY, Modified, Jackson, Size 8	I
6515-00-518-5498	CANNULA, TRACHEOSTOMY, Modified, Jackson Size 8, with 15 mm adapter	II

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

Preparing Activity

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4836

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54271	CANNULA, TRACHEOSTOMY, MODIFIED, JACKSON

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI – Optional)*	b. WORK TELEPHONE NUMBER *(Include Area Code – Optional)*
c. MAILING ADDRESS *(Street, City, State, ZIP Code – Optional)*	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.