COMMERCIAL ITEM DESCRIPTION

FORCEPS, ENDOSCOPIC, Globular Object Grasping (Jackson)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a globular object grasping endoscopic forceps of the Jackson design, suitable for use in endoscopic procedures.

Salient characteristics:

General. Shall be a cannulated endoscopic forceps of the Jackson design, with a 60 cm working length, having the following size and style:

Size 1 style 1 shall fit through endoscopes 6 mm or larger and shall have cup-shaped, fenestrated jaws. As an alternate for size 1 style 1, the jaws may not be fenestrated.

Size 2 style 2 shall fit through endoscopes 9 mm or larger and shall have disc-shaped, fenestrated jaws.

The box lock instrument shall have jaws which are shaped and set to provide a secure grip on the globular object being removed.

Material. The jaws and box lock portion shall be fabricated from Class 4 martensitic corrosion-resistant steel in accordance with ASTM Designation F 899. The cannulated tube shall be fabricated from Class 4 steel or Class 3 austenitic stainless steel in accordance with ASTM Designation F899.

Hardness. The jaws shall have a Rockwell Hardness C scale (HRC) of 48 to 53 in accordance with ASTM Designations E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Designation E140. The hardness of the tube shall be Rockwell B75 minimum. The remainder of the instrument shall be Rockwell C40 to C47.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil --Downloaded: 202640-26T09:32Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing #22358.

Passivation. The forceps shall be passivated.

Finish. The forceps shall be mirror, satin, or matte. The matte finish shall exhibit a uniform, non-directional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Corrosion-resistance. The forceps shall show no signs of corrosion or plating of copper when tested for corrosion-resistance in accordance with ASTM Designation F 1089.

Performance. When fully closed, the jaws shall align with each other (no scissor). Action of the forceps shall be easy, without drag or binding. Performance shall be in accordance with applicable portions of ASTM F 1026.

Markings. The forceps shall bear the contractor's name or registered trademark and the word "stainless". Forceps of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The forceps shall be free from burrs, cracks, sharp edges, scale, corrosion or other defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Sampling for inspection shall be level S-2. Dimensional testing shall be at a 2.5% AQL, hardness testing at a 1.5% AQL, and corrosion-resistance testing at a 1.0% AQL. Unit of product for test purposes shall be one forceps.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Numbers are covered by this document:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION	SIZE	STYLE
6515-01-139-8381	FORCEPS, ENDOSCOPIC, Medium, Globular Object, Grasping, Jackson, 60 cm Long	1	1
6515-01-139-8382	FORCEPS, ENDOSCOPIC, Large, Globular Object, Grasping, Jackson, 60 cm Long	2	2

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4462

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:32Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54276	FORCEPS, ENDOSCOPIC, Globular Object Grasping (Jackson)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional	b WORK TELEPHONE NUMBER (Include Area Code) – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:32Z
Check the source to verify that this is the current version before use.