A-A-54281
3 December 1990

COMMERCIAL ITEM DESCRIPTION

HOLDER, SUTURE NEEDLE (4-5/8 INCH, TROUTMAN)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a 4-5/8 inch long suture needle holder, of the Troutman design, suitable for use in microsurgery

Salient Characteristics:

General. The suture needle holder shall have delicate jaws. The jaws shall be curved, flattened and shall have a wide flat handle and a lock

Material The suture needle holder shall be fabricated from Class 4 martensitic steel in accordance with ASTM Designation F899.

Style, design and dimensions. The suture needle holder shall be be 4-5/8 inches long with serrated handles to limit slippage. The jaws faces shall not be serrated. When closed, there shall be no protruding points. The jaws shall close, tips first, before locking. The suture needle holder shall be in accordance with Defense Personnel Support Center (DSPC) Drawing 23405

Hardness The suture needle holder shall have a Rockwell Hardness C scale (HRC) of 40 to 47 in accordance with ASTM Designation E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM Designation E140.

Finish. The suture needle holder shall have a mirror and satin finish

Passivation The finished instrument shall be properly passivated in accordance with ASTM Designation F1026

Corrosion resistance The suture needle holder shall show no signs of corrosion or plating in accordance with ASTM Designation F1089

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Identification marking and labeling The suture needle holder shall bear the manufacturer's name or registered trademark and the word 'stainless' Holders of foreign manufacture shall bear the name of the country of origin

Workmanship The suture needle holder shall be free from sharp edges, burrs, cracks, grind marks or other defects which detract from its appearance or impair its serviceability

Unit Each (EA) One holder, as specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Tests Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein The inspection level shall be S-2 Dimensional testing shall be at a 2 5 % AQL, hardness testing at a 1 5% AQL Corrosion and performance testing shall be at a 1 0% AQL The unit of product for tests purposes shall be one suture needle holder

Holder performance Clamp a 0 005 ± 0 001 inch o d cotton or silk suture (common thread may be substituted if dimensional and tensile strength requirements are met) in the jaws and subject the suture to a 0.45 pound force (minimum) for 30 seconds minimum Any slippage of the suture in the jaws shall constitute a test failure Clamp a 0 020 +\- 0 004 inch o d suture needle (or, as an alternate a steel wire of the same dimensions) in the needle holder jaws The needle shall be clamped perpendicular to the axis of the needle holder such that a portion of the needle is exposed on both sides of the jaws Apply one pound force such that the force is distributed evenly on the metal protruding on either side of the jaws Any slippage of the needle in the jaws resulting in a displacement of the needle (i e. side to side twist is excluded) shall be a test failure

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-098-5806	HOLDER, SUTURE NEEDLE, TROUTMAN, 4-5/8 Inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No 6515-4069

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN DPSC-BSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54281	HOLDER, SUTURE NEEDLE (4-5/8 INCH, TROUTMAN)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) - Optional	8. DATE OF SUBMISSION (YYMMDD)

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.