A-A-54286
23 November 1990

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

(Universal)

The General Services Administration has authorized the use of this commercial item description

This Commercial Item Description covers a sterile disposable, universal surgical drape pack for use in creating a sterile field for any surgical procedure

Salient Characteristics

The surgical pack shall be a sterile, disposable universal surgical drape pack fabricated of a nonwoven material for creating a sterile field for any surgical procedure The pack shall consist of at least the following components

a One overwrap/table cover 50 by 90 inches
b One Mayo stand cover with pad, 23 by 55 inches, plus or minus 1 inch
c One scrub nurse gown with towel, overwrapped
d Two extra protection surgical gowns with towels
e One absorbent towel, 19 by 19 inches, plus or minus 1 inch
f Ten feet of nonconductive tubing
g One Yankauer suction tip
h One suture bag
i One bottom drape 72 by 70 inches, plus or minus 2 inches
j One top drape 100 by 57 inches, plus or minus 2 inches
k Two side drapes, 36 by 77 inches, plus or minus 2 inches

The overwrap/table cover shall be far-folded for ease in opening on the surgical room table The overwrap shall be impervious to fluids and moisture

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The Mayo tray and stand cover shall be fabricated of a plastic material at least 2 75 mils thick and shall be impervious to fluids The cover shall be fan-folded with a six inch cuff on the distal end for sterile draping of the Mayo tray

The gowns shall be large or extra-large size All gowns shall

a Be soft, comfortable, pliable, quiet and memory free
b Have high breathability, facilitating comfort for wearing for long periods of time.
c Be impervious to fluid The two surgical gowns shall have extra protection against fluid penetration
d Have stockinette cuffs that shall conform to the wrists
e Be wrap-around type to completely cover the back of the individual gowned
f Have a tie closure at the waist that shall not rip or separate when tying The gowns shall be designed to minimize the possibility of contamination when tying

The drapes shall each have an adhesive strip along one edge for attachment to the patient, a bacterial fabric reinforcement on each drape and cord holders The bottom drape shall have clip tabs The reinforced area shall be impervious to fluids yet highly absorbent eliminating strike-through while preventing pooling of fluids

All nonwoven items shall be fabricated of a soft material with minimal memory for good drapeability The nonwovens shall resist tearing, pilling and linting

If pack is opened and not used it shall be capable of sterilization by steam at 275° F or by ethylene oxide

The pack shall be sterile and individually wrapped in a suitable tear-resistant, waterproof heat sealed package which shall allow aseptic handling and maintain sterility of the contents until opened or damaged The items shall be packed in sequence of use

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

All pack components shall be clearly listed on the exterior wrapper which shall also serve as a dust cover

Workmanship The surgical packs shall be free from defects which detract from their appearance or impair their serviceability

Unit Package (PG) One package containing three individually packaged, sterile surgical packs, as specified, constitutes one unit The individual package shall be a tear resistant, waterproof heat sealed package, which shall allow aseptic handling and shall be capable of maintaining sterility of contents unless the package is damaged or opened Items shall be packed in sequence of use

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection as used herein is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one surgical pack

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2 5 (percent defective) and an inspection level of S-1 The unit of product for test purposes shall be one surgical pack

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include, but shall not be limited to, the following

Sterility test The surgical packs shall be testec for sterility in accordance with the recuirements of the Food and Drug Administration

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standarc 376, and all other requirements of this cocument are met

If a product is manufacturec tc metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units a request shoulc be macie to the contracting officer to cetermine if the procuct is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offerec meets the salient characteristics of this cescription anc conforms tc the producers' own crawings specifications, stancaros, and quality assurance practices The Government reserves the right to require proof of such conformance prior tc first ceiivery and thereafter as may be otherwise provicec for under the provisions of the contract

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE The following National Stock Number is covered by this document

6530-01-305-6987

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

A-A-54286

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
USPHS
FDA-MPQAS

Project Number 6530-2184

Location PACKCID 15E

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.