COMMERCIAL ITEM DESCRIPTION

STOPPER, BOTTLE (White Rubber)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-36662 which is cancelled.

This Commercial Item Description covers perforated bottle stoppers with removable plugs, for use as solid or perforated bottle stoppers in a chemical or clinical laboratory.

Salient Characteristics:

Shall be white color rubber stoppers for use as closures on bottles used in laboratory procedures.

Shall be suitable for use as both a solid stopper and a perforated stopper.

Shall be fabricated from natural rubber or synthetic rubber or from a mixture of natural and synthetic rubber.

Shall be supplied with full length perforations (holes) that are countersunk on the top surface, and sealed on the bottom surface with protruding and removable nodules (plugs). Each perforation (hole) shall have a millimeter (mm) diameter of 5 mm +/- 0.5 mm. The number of perforations (holes) for each specification designated size shall be as hereinafter indicated.

Shall be of the following stopper number sizes with the indicated dimensions and number of perforations (holes):

DIMENSIONS OF STOPPERS (MM)

Stopper Number Size	Number of Holes	Diameter of Top	Diameter of Bottom	Length
3	2	24	18	25
4	2	26	20	25
5	3	27	23	25

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

DIMENSIONS OF STOPPERS (MM)

Stopper Number Size	Number of Holes	Diameter of Top	Diameter of Bottom	Length
6	3	32	26	25
7	3	37	30	25
8	3	41	33	25
9	3	45	37	25
10	3	50	42	25

A +/- 1 mm tolerance shall be applicable to the required dimensions for the diameters (top and bottom) and lengths of the stoppers.

Shall have each stopper supplied with its stopper number size identified in the form of a raised number on the top side.

Shall have the bottom side of each perforation (in the stopper) supplied with an inserted protruding nodule (plug) that can be removed by twisting off with finger pressure The removal of each nodule (plug) shall result in an unobstructed full length opening that permits the full insertion of laboratory glass tubing.

Shall be supplied in unit packages (see 'Unit' requirements) that include the following legibly printed legend 'CAUTION- Stoppers are rubber based and may become tacky when used with organic solvents'

Workmanship. The bottle stoppers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). For stopper numbers 3 and 4,one package containing fifty bottle stoppers, as specified, constitutes one unit. For stopper number 5, one package containing forty bottle stoppers, as specified, constitutes one unit For stopper number 6, one package containing thirty bottle stoppers, as specified, constitutes one unit. For stopper numbers 7 and 8, one package containing twenty bottle stoppers, as specified, constitutes one unit. For stopper number 9, one package containing fifteen bottle stoppers, as specified, constitutes one unit. For stopper number 10, one package containing ten bottle stoppers, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the bottle stoppers in accordance with commercial practice Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Stopper number(s) required, intermediate/ exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Numbers are covered by this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

NSN	Item Identification	Stopper Number
6640-00-935-1102	STOPPER, BOTTLE, White Rubber, Solid or Two Hole, No 3, 50's	3
6640-00-935-1103	STOPPER, BOTTLE, White Rubber, Solid or Two Hole, No 4, 50's	4
6640-00-935-1104	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No. 5, 40's	5
6640-00-935-1105	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No. 6, 30's	6
6640-00-935-1106	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No 7, 20's	7
6640-00-935-1107	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No. 8, 20's	8
6640-00-935-1108	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No. 9, 15's	9
6640-00-935-1109	STOPPER, BOTTLE, White Rubber, Solid or Three Hole, No. 10, 10's	10

NOTE: Commercial one pound bags of bottle stoppers, conforming to specification requirements, are acceptable provided the quantity of bottle stoppers within the bag includes (as a minimum) the quantity required by the applicable National Stock Number.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1637

Location: ENABLE/STOPERC2/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. *If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.*

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54293	STOPPER, BOTTLE (White Rubber)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
8. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	9. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.