A-A-54294
14 December 1990

COMMERCIAL ITEM DESCRIPTION

INSERT, HANDPIECE, DENTAL ULTRASONIC
PROPHYLAXIS UNIT

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers replacement inserts for dental ultrasonic prophylaxis unit 3000.

Salient characteristics:

The handpiece inserts shall be suitable for use with dental ultrasonic prophylaxis unit 3000 (see Notes).

The replacement inserts shall be as specified below:

Style I - 30K-3, Flat-Hoe Chisel Type
Style II - 30K-10, Curette Type
Style III - 30K-EWPP, Perio Probe Type

Style I. The insert shall be used for removing supragingival calculus from labial and lingual surfaces of teeth and removing surface stain.

Style II. The insert shall be suitable for use on all supragingival and shallow subgingival tooth surfaces.

Style III. The insert shall be used for subgingival scaling and root planing.

The inserts shall include an anti-rotation key to prevent rotation of the tip. The operative tip shall be in line with the central unit of the handle. The flow through the groove in the tip, working in conjunction with the metal eyelet at the juncture of the tip and nozzle shall assure maximum cavitation of water at the tip with a minimal formation of water droplets at the juncture of the tip and the nozzle.

The tip shall be activated by mechanical oscillations of the strips which make up the transducer. These oscillations shall result in elliptical tip motions at a frequency of 30,000 tip motions per second over a distance of 0.003 inch.

The inserts shall be autoclavable.

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Material.

Tip. The tip shall be fabricated of type 420 martensitic stainless steel, as specified in ASTM Designation F899, for composition only.

Connecting body. The connecting body shall be fabricated of Monel 400.

Nozzle. The nozzle shall be fabricated of polysulfone.

Transducer (magnetostrictive stack). The transducer shall be fabricated of Permanickel Alloy 300 strips. Each strip shall be 0.0095 to 0.0105 inch thick and 0.178 to 0.187 inch wide. The thickness of the transducer shall be 0.159 to 0.168 inch thick. The length of the transducer shall be 3-3/4 ± 1/16 inches.

Hardness. The operative tip shall have a Rockwell hardness of C40 to C45 as specified in ASTM Designation E18 (Rockwell) or E92 (Vickers) and E140.

Finish. The operative tips shall have a bright finish.

Identification marking. Each insert shall be legibly and permanently marked with the manufacturer's name or trademark.

Workmanship. The nozzle, transducer and tips shall show no evidence of cracking, holes, pitting, metal slivers, burrs, sharp edges or other defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One insert of one style only, as specified, constitutes one unit. Each insert shall be individually packaged in a blister type plastic.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, test purposes shall be one insert of each type unless otherwise specified, sampling for examination shall conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0 (percent defective).

Examination. Examination shall include but shall not be limited to the following defects:

Defects.

Anti-rotation key missing or does not prevent tip rotation.
Insert does not fit properly into handpiece
Identification markings not legible, missing, or incomplete
Workmanship not as specified

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Chemical analysis. In the absence of certificates of quality for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Contract certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirement.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-281-8793	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT, 30K-10, Curette Type
6520-01-313-6269	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT, 30K-3, Flat-Hoe Chisel Type
6520-01-313-6270	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT, 30K-EWPP, Probe Type

NOTE: The following National Stock Number is referenced by this document.

National Stock Number	Item Identification
6520-01-313-6250	ULTRASONIC PROPHYLAXIS UNIT, DENTAL, PORTABLE, 30K

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2400

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-MSTE
2800 South 20th Street

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54294	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.