INCH-POUND

A-A-54295A
31 May 1993
SUPERSEDING
A-A-54295
1 April 1991

COMMERCIAL ITEM DESCRIPTION

TUBE, SUCTION, SURGICAL, PLASTIC, Disposable

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a one type of plastic surgical suction tube suitable for suctioning procedures where a sterile tube is required.

Salient characteristics:

General. The suction tube shall be sterile, disposable and non-conductive. The tube shall be used for applying a vacuum to suction instruments and other devices, which remove fluids and gases from a patient during and after surgery to a collection device.

Material. The suction tube shall consist of clear, plastic, nonconductive tubing. The tube shall have a minimal coil memory for good drapeability in sterility fields.

Style and design. The suction tube shall have molded female connector ends and a plastic male connector in order to permit connection to other tubing. The tubing connectors shall make mechanically secure and leakproof connection with other tubing connections.

The suction tube shall meet these additional requirements:

Inside Diameter - 9/32 ± 1/64 inches (7.0 mm ± 0.4 mm)
Nominal Length - 12 feet
Minimum Breaking Strength - 25 pounds
Minimum Ultimate Elongation - 100 percent
Collapsibility - shall not collapse in vacuum of 22 inches of mercury applied for 24 hours
Biocompatability - Class IV plastic of the United States Pharmacopeia (USP) applies

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Connectors bonding strength. The bonded connectors shall withstand a pull force of 10 lbs without separation from the tubing.

Sterility. The suction tube shall be sterile in accordance with the USP or AAMI guidelines. Sterility assurance level shall be 0.0001.

Workmanship. The suction tube shall be free from defects which detract from its appearance or impair its serviceability. The tube shall be free of cuts, rough areas, flash, weak areas, discoloration and foreign matter.

Unit. Package (PG). One package containing 20 individually packaged sterile suction tubes as specified, constitutes one unit. Tubing shall be housed in an inner package of a type to permit aseptic handling when outer package is opened. The individual package shall be a sealed peel-open container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:32Z
Check the source to verify that this is the currentversion before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one tube as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one tube. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:32Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-174-2354	TUBE, SUCTION, SURGICAL, Plastic, Disposable, 9/32 inch (7.1 mm) Inside Diameter

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:32Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5267
Location: c:\word\doc\nish\tubecid\sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.