A-A-51025
3 October 1985

COMMERCIAL ITEM DESCRIPTION

SWAB, PROCTOSIGMOIDOSCOPIC TIP, (Non-Sterile, 500's)

The General Services Administration has approved the use of this commercial item description.

This commercial item description covers a proctosigmoidoscopic swab suitable for use in sigmoidoscopic procedures. The swab shall conform to the salient characteristics listed below.

Salient characteristics:

Material.

Fibrous tuft. The tuft of the swab shall be fabricated from white, USP purified rayon fibers. In addition, the purified rayon fibers shall have a lead content that does not exceed 5 parts per million (ppm) when tested in accordance with the USP.

Stick. The stick shall be fabricated from white sulfate converting paper sized with methyl cellulose and its moisture content shall be not greater than 6 percent by weight when tested in accordance with ASTM Test Method D2044-68.

The swabs shall be provided clean and ready for use.

Construction. The swab shall consist of a stick component and a fibrous tuft rayon component. The paper material used to fabricate the stick shall be tightly rolled into a uniform cylinder. One end of the stick shall be tightly wrapped with a pear-shaped fibrous tuft rayon tip. Swab sticks shall be straight and free from fraying or stray filaments. The tip shall not unravel, loosen or become detached when exposed to water or a humid atmosphere. The swab shall be flexible and shall not break when bent through an arc of $145^{\circ} \pm 5^{\circ}$.

Dimensions. The swab shall have an overall length of 16 ± 1 inch (40.6 ± 2.5 cm). The diameter of the stick shall be 5/32 ± 1/32 inch (4.0 ± 0.79 mm). The maximum diameter of the tip shall be 15/32 ± 1/32 inch (11.0 ± 0.79 mm).

Workmanship. The swabs shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing five sealed bags, 100 swabs per bag, constitutes one unit. (500 swabs)

9/12/85
FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or processes) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one swab.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-01-163-8077	SWAB, PROCTOSIGMOIDOSCOPIC TIP, Non Sterile, 500's.

MILITARY INTERESTS

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY.

DoD-MB

Agent

DLA-DM

52250

Project No. 6515-3283

S1-8-3

*U.S. GOVERNMENT PRINTING OFFICE 1988 — 605 036/24802

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.