A-A-51034
10 October 1985

# COMMERCIAL ITEM DESCRIPTION

## DISPENSER, SURGICAL DETERGENT

This commercial item description shall be used in preference to Federal Specification No. FF-D-400.

This commercial item description covers a liquid surgical detergent and liquid soap dispenser for use in military medical treatment facilities.

Salient characteristics:

Shall be a wall mounted, foot operated, liquid surgical detergent and liquid soap dispenser.

Shall consist of a dispensing nozzle, a metering device, a bowl or jar, a foot operated dispenser activating device with interconnecting flexible plastic tubing, and a bowl or jar retainer adapted for wall mounting.

Shall have a dispensing nozzle constructed from suitable corrosion-resisting steel or nickel chromium alloy.

Shall have a sturdy plastic bowl or jar with a minimum wall thickness of 3/32 inch and a nominal capacity of one quart.

Shall have all dispenser components constructed from suitable materials to resist corrosion without contaminating, discoloring or adversely affecting the detergent or soap.

Shall have liquid flow control devices, ports and passages designed to avoid clogging.

Shall have a dispenser capable of delivering a minimum of 1.9 milliliters and a maximum of 3 milliliters for each full depression of the foot actuating device. The flow of liquid shall be controlled by either a positive displacement type or an adjustable air release type valve. The quantity of detergent or soap delivered shall be factory adjustable between these limits.

Plastic tubing connecting the foot-pedal accuator to the dispensing mechanism shall have minimum dimensions of 1/16 inch wall thickness and 5 feet, 6 inches length.

Dispenser shall be free of burrs and sharp or ragged edges. All plastic shall be free of mottling, bubbles, cracks, pitting, and flash. Metal shall be polished to a uniform commercial finish.

Shall include service data including instructions for use and maintenance.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:49Z
Check the source to verify that this is the current version before use.

Workmanship. The detergent dispenser shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). Each dispenser, as specified, constitutes one unit.

ontractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-299-8114

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 202640-26T08:49Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 50628

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6530-1774

Location H1-11-4

U S GOVERNMENT PRINTING OFFICE: 1985 - 605-036/24839



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.