A-A-51051
7 October 1985

COMMERCIAL ITEM DESCRIPTION

GUIDE WIRE, CARDIOVASCULAR CATHETER, Fixed Core, Heparin Coated

The General Service Administration has authorized the use of this commercial item description.

This commercial item description covers the requirements for a guide wire suitable for use during cardiac catheterization procedures.

Salient characteristics:

Shall be a sterile, straight, fixed core guide wire, 0.035 inch (0.89 cm) in diameter and 145 cm long.

The guide shall be constructed of tightly spiraled wire over the fixed core surface. The end of the guide shall have a 3 cm flexible tip.

A microthin polytetrafluoroethylene coating shall be applied to the fixed core to reduce friction in the catheter. The guide wire shall have a bonded heparin coating.

The entire guide wire shall be protected within a plastic guide holder. The holder shall be color coded by guide diameter facilitating identification and accurate handling during insertion of the guide.

The guide wire shall be nontoxic and pyrogen free and shall meet any Federal Food and Drug Administration (FDA) requirements for safety and toxicity applicable for invasive and implantable devices.

Each wire guide shall be individually wrapped in a suitable, tear-resistant peel-open package which will allow aseptic handling and maintain sterility of the contents until opened or damaged. A sterility indicator shall be included in each package. The sterility of the guide wire shall be tested in accordance with the USP.

The guide wire with the plastic holder shall be capable of withstanding ethylene oxide sterilization. The guide wire without the plastic holder shall withstand steam sterilization.

The guide wire shall be fabricated from type 302 austenitic corrosion-resisting steel.

Workmanship. The guide wire shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One guide wire, as specified, constitutes one unit.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one guide wire.

Contractor certification. The contractor shall certify that the product offered meet the salient characteristics of this description and conforms to the producers' own drawing, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-166-5111

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS.

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS

Project No. 6515-3394

51013

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.