A-A-51053
7 October 1985

COMMERCIAL ITEM DESCRIPTION

TUBE, STOMACH, SURGICAL,
Silicone, Salem, Disposable

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers sterile, disposable, radiopaque, double-lumen stomach tubes of the Salem design, suitable for long term enteral feeding and aspiration of the stomach contents.

Salient characteristics.

General. The salem, surgical, stomach tube shall be suitable for long term enteral feeding and aspiration of the stomach contents.

Characteristics. The sterile, disposable double-lumen, Salem, surgical stomach tube shall have a radiopaque line running lengthwise within its wall. The stomach tube shall have a length of 48 inches. The stomach tube shall be 16 Fr. The stomach tube shall be rounded and closed at the distal end.

There shall be ten oval eyes at the distal end. The ten oval eyes shall be equally distributed on each side of the radiopaque stripe. Another oval eye shall bisect the radiopaque stripe not less than 2 inches from the distal end.

The proximal end shall have a permanently attached connection to the two lumens of the tube. Within the wall of the main lumen there shall be a secondary lumen. The main lumen shall function as a suction drainage lumen. The secondary lumen shall serve as an air vent and can also be used for irrigation. The proximal end of the stomach tube shall be flared to accommodate a removable universal size connector. This connector shall be provided with the stomach tube. A flanged suction lumen adapter, with cap, shall be provided for the main lumen. The lumen adapter shall facilitate connection to catheter tip syringes. A lumen cap shall be provided for the secondary lumen. The stomach tube shall be marked at 18, 22, 26 and 30 inches from the distal end of the tube.

Radiopacity. Select at least two tubes from each lot and radiograph as follows.

a) Superimpose 240 ± 10 mm of USP paraffin over the tubes.
b) Use a standard make of cassette with screens, films and processing solutions.
c) Expose at a setting of 85 kV, 10mAs and from a distance of 640 ± 10 mm.

The tubes shall have a stripe that shows as a definite shadow when tested.

Sterility. Each stomach tube shall be sterile when tested in accordance with USP.

Safety. The stomach tubes shall be tested for safety in accordance with the USP for transfusion and infusion assemblies.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Material. The stomach tube shall be fabricated from medical grade silicone. The connector shall be fabricated from plastic.

Testing. The stomach tube shall be in accordance with the requirements of USP Class 6 material.

Workmanship. The stomach tube shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten tubes, each tube individually packaged sterile in a peel-open type container, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be a stomach tube.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number is covered by this document

6515-01-201-6998

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY.

DoD - MB

Agent

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY.

VA - OSS

58315

Project No. 6515-3421

Location S1-9-14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.