COMMERCIAL ITEM DESCRIPTION

GRADUATE, LIQUID, LABORATORY, (Corrosion Resisting)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-G-36614.

This commercial item description covers one type of corrosion resisting laboratory graduate container.

Salient characteristics:

The Type I graduate shall be supplied without a handle, and the Type II shall be supplied with a handle. The Size 1 shall have a capacity of 500 milliliters (ml), and the Size 2 graduate shall have a capacity of 1000 ml.

Table I.

Item Identification	Specification Designation	
	Type	Size
GRADUATE, LIQUID, LABORATORY, Corrosion-Resisting Metal, 500ml.	I	1
GRADUATE, LIQUID, LABORATORY, Corrosion-Resisting Metal, 1000ml.	II	2

The graduates shall have a minimum thickness of .014 inches (28 Gauge) and .013 (29 Gauge) for Type I and II respectively, fabricated from corrosion resisting steel. The graduates shall be uniform in construction and free from surface imperfections.

The bottoms of the type I and II graduates shall be circular and perpendicular to the vertical axis. Graduate bottoms that are concave shall be acceptable. The sides of the graduates shall taper upward from the bottoms (i.e. away from the vertical axis towards the top edge). The upper edges of the graduates shall be flared outward and shall have a pour-out spout. For the type II graduate, a handle shall be attached to the body at the end opposite the spout. The handle shall be attached by spot welding or resistance welding. Welds shall be as strong as the parent material.

The nominal dimensions for the overall height and base diameter of the size 1 graduate shall be 4-9/16 and 2-7/8 inches respectively. The width of the top edge along the symmetrical axis shall be 3-5/8 inches. The nominal dimensions for the overall height and base diameter of the size 2 graduate shall be 5-1/2 and 3-13/16 inches respectively. The width of the top edge along the symmetrical axis shall be 4-13/16 inches. All dimensions shall be $\pm$ 1/16 inch.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The insides of the type I and type II graduates shall be graduated in a double scale with markings that are permanent and legible. One scale shall be in ounces (oz) or fluid ounces (fl oz), and the other scale shall be in milliliters (ml) or cubic centimeters (cc). Grams (gm) may be added to the metric scale. For the size 1 graduate, the oz (fl oz) scale shall be graduated from 2 to 16 ounces in increments of 2 ounces. The metric scale shall be graduated from 50 to 500 ml in increments of 50 ml. For the size 2 graduate, the oz scale shall be graduated from 4 to 32 oz in increments of 4 oz. The metric scale shall be graduated from 100 to 1000 ml (cc) in increments of 100 ml (cc).

The graduation markings for the size 1 and size 2 graduates shall be accurate to within ± 5 percent for greater than 100 ml.

Workmanship. The container shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One graduate, of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be conducted at inspection level S-II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-II, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-427-5025	GRADUATE, LIQUID, LABORATORY Type I, Size 1
6640-00-431-4520	GRADUATE, LIQUID, LABORATORY Type II, Size 2

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51059

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Aie Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6640-1325

Location S3-7-8

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 — 605-036/45066

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51059	GRADUATE, LIQUID, LABORATORY,

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.