A-A-51065
10 October 1985

COMMERCIAL ITEM DESCRIPTION

TUBE, SUCTION, SURGICAL
Rhoton-Merz, Straight Tip

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers surgical suction tubes of the Rhoton-Merz design in French sizes of 3, 5 and 7.

Salient characteristics:

The tubing shall be a slightly malleable metal suction tube of 8 cm in length. The distal portion of the suction tube shall be straight with slightly enlarged tip to minimize tissue damage. The tubing shall have a finger valve to control the amount of suctioning.

The suction tube shall be angulated at the area just distal to the finger control valve to facilitate visibility during suctioning. The French size shall be clearly marked on the finger control valve. The proximal end shall be grooved to facilitate tubing attachment.

The surgical tube shall be fabricated of a suitable corrosion-resisting steel, properly hardened and passivated to maintain slight malleability.

Workmanship. The suction tube shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One suction tube, as the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein.. The unit of product for test purposes shall be one suction tube.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers (NSNs) and their respective French sizes are covered by this document

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

NSN	French Size
6515-01-164-8814	3
6515-01-164-8815	5
6515-01-164-8816	7

Custodians

Army-MD
Navy-MS
Air Force-03

Preparing Activity

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITY

VA-OSS

Agent

DLA-DH

06737

Project No. 6515-3405

Location S1-9-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.