A-A-51070A
30 August 1991
SUPERSEDING
A-A-51070
25 October 1985

COMMERCIAL ITEM DESCRIPTION

MASK, SURGICAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Type III of MIL-M-36954 which is cancelled.

This Commercial Item Description covers a disposable, cupped, nonwoven fabric surgical mask.

Salient Characteristics:

The surgical mask shall be disposable, cupped, lightweight, fluid resistant and suitable for use by personnel in operating rooms and by other personnel in maintaining protective aseptic techniques.

The body of the mask shall be formed into a cup shape, the periphery of which shall be contoured to fit over the bridge of the nose, under the eyes, along the cheeks, and under the chin. It shall be formed of a porous, nonwoven, fibrous fabric treated with a suitable, water-resistant sizing agent. The mask body shall serve as the filter and a fluid barrier. The characteristics of the formed coated fabric shall be such that the mask body will retain its cup shape under normal handling and use, but will permit adjustment, and conformance to the contours of the face. The mask body shall be blue or a suitable color with a matte finish.

The area of the body which fits over the bridge of the nose shall have a malleable metal strip or nosepiece suitably cemented to the outside surface near the edge. This nose-piece shall be capable of being formed easily to the contours of the wearer's nose, and maintaining these contours

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

to minimize the loss of expired air between the mask and face in this area.

The cupping of the body shall be sufficiently deep that the wearer's nose and mouth do not touch the mask. The portion of the cup over the nose and mouth shall be ribbed or corrugated to enhance the rigidity in this area.

A suitable elastic headband shall be permanently attached to the body at its two sides, the points of attachment shall be located so that the band will hold the mask firmly against the face. The length of the unstretched band between attachment points shall be at least 13 inches.

The porosity of the mask material shall be such that:

* The pressure differential across a sample shall not exceed 0.09 inches of water when a 17.5 square inch specimen from the center of the mask is exposed to an enclosed air stream of 50 plus or minus 2 liters per minute; and

* The filtering efficiency of the mask shall be not less than 95 percent when tested by the Greene and Vesley (J. Bacteriol. Vol. 83, No. 3, P-663, March 1962) or a comparable method.

The mask shall meet the requirements of Class 1 Federal Regulation CFR 1610 Part 16. The mask shall not cause irritation to the skin and eyes when tested using human panel and rabbit subjects in compliance with Federal Regulation CFR 16 1500.41 and 1500.42.

The mask shall be fluid resistant as defined by meeting the following requirement when tested in accordance with Hydrostatic Pressure Test Method AATCC 127-1985:

Minimum average = 6.0 cm
Sample size = 15

Workmanship. The masks shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 50 masks, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one mask.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one mask.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Porosity, filtering efficiency, flammability, skin irritation, and fluid resistance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-00-982-7493

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5106

Location: MASK3CID.15E

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-51070A	2. DOCUMENT DATE (YYMMDD) 30 August 1991

3. DOCUMENT TITLE
MASK, SURGICAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial AC 215-737-2870	(2) AUTOVON AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.