W-L-316B
September 22, 1976
SUPERSEDING
Int. Fed. Spec. W-L-00316a (DSA-DM)
January 23, 1967 and
Fed Spec. W-L-316
August 21, 1958

FEDERAL SPECIFICATION

LIGHT, FLOOR

This specification was approved by the Commissioner, Federal Supply Services, General Services Administration, for use of all Federal agencies.

1 SCOPE

1.1 This specification covers a portable, floor standing, operating and examining light, capable of accommodating a standard incandescent lamp bulb

2 APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein

Federal Specifications

QQ-C-320	-	Chromium Plating (Electrodeposited)
QQ-N-290	-	Nickel Plating (Electrodeposited)
PPP-B-585	-	Boxes, Wood, Wirebound
PPP-B-601	-	Boxes, Wood, Cleated-Plywood
PPP-B-621	-	Boxes, Wood, Nailed and Lock-Corner
PPP-B-636	-	Boxes, Shipping, Fiberboard.

Federal Standards

Fed Std No. 123 - Marking for Shipment (Civil Agencies)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Forth Worth, Denver, San Francisco, Los Angeles, and Seattle, WA

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies

Military Specifications

MIL-A-8625 - Anodic Coatings for Aluminum and Aluminum Alloys
MIL-L-10547 - Liners, Case, and Sheet, Overwrap, Water-vaporproof, or Waterproof, Flexible.

Military Standards

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications, and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications The following documents form a part of this specification to the extent specified herein Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the currentversion before use.

Underwriter's Laboratories, Inc. (UL):

UL 153 Portable Electric Lamps
UL 498 Electrical Attachment Plugs and Receptacles
UL 544 Medical and Dental Equipment

(Application for copies should be addressed to the Underwriter's Laboratories, Inc 207 East Ohio Street, Chicago, IL 60611, 333 Pfingsten Rd , Northbrook, IL 60062, 1285 Walt Whitman Rd , Melville, L I., NY 11746, 1655 Scott Blvd., Santa Clara, CA 95050.)

American National Standards Institute (ANSI) Standard

C73 11 Dimensions of Attachment Plugs and Receptacles 125 Volts, 15 Amperes, 2 Pole, 3 Wire, Grounding Type, General Purpose

(Application for copies should be addressed to the American National Standards Institute Inc., 1430 Broadway, New York, NY 10018)

3 REQUIREMENTS

3 1 First article When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4 1.2 1 and 6 2)

3 2 Fire and casualty hazards

3 2 1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of Underwriter's Laboratories (UL) Inc , Standards of Safety, UL 153, UL 498, and UL 544 Proof shall consist of one of the following

A UL label affixed to the product, provided the contractor agrees to have Underwriters' Laboratories, Inc forward a copy of the UL report directly to the Contracting Officer, upon the latter's request

B Laboratory report from a nationally recognized independent laboratory If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards The Government reserves the right to require a facilities report and to conduct a facilities survey Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches, and parts list, all appropriately identified. Compliance with this paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.2.2 Compliance with the above requirements regarding fire and casualty hazards does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.

3 3 Materials All material used in the construction of the light shall be of the highest quality throughout, suitable in every respect for the purpose intended Standard commercial tolerances shall apply to thickness of steel sheet and to thickness and diameter of tubing

3 3 1 Tubing Tubing shall be brass or steel, seamless or welded, mechanical tubing.

3.3.2 Base. Base shall be cast iron or pressed steel.

3.3 3 Base housing Sheet used in the base housing shall be carbon steel

3.3 4 Reflector The reflector shall be carbon steel, 024 inch thick (min) or aluminum. .030 inch thick (min.)

3 4 Design and construction The light shall be in accordance with figure 1 The design shall permit continuous height adjustment, measured from the floor to the top of the telescoping tube, of not more than 35 inches at the lowest position and not less than 54 inches at the highest position

3 4 1 Base The base shall be made of cast iron or pressed steel. Cast iron base shall have a base housing of 0239 inch thick sheet steel. Pressed steel base may be weighted with cast iron or lead for stability. The base shall be round, not less than 10 inches nor more than 11 1/2 inches in diameter The top of the base shall be not more than 2 inches above the floor and the upper edge shall be rounded, with an outer radius of not less than 2 inches The light shall rest evenly on a flat surface without rocking. The base shall not have any burrs, slivers, or sharp edges. Metal portions of the base in contact with the floor shall be ground smooth to prevent scratching the floor surface. The full circumference of the base shall rest upon the floor. Alternatively, the base shall rest upon three metal embossments or upon three round, conductive rubber pads, spaced evenly around the circumference. Rubber

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

pads shall be at least one inch in diameter, shall extend at least 1/4 inch below the base, and shall be secured with screws so that they can be removed and replaced Metal embossments or rubber pads shall not extend beyond the circumference of the base Optionally, extra metal embossments, not normally in contact with the floor, may be spaced between the pads upon which the light rests to add stability against tipping The base shall be so configured, and sufficiently heavy, that the light shall not tip over when tested as specified in 4 4 1.

3 4.2 Upright housing and telescoping tube The upright housing shall be 7/8 inch outside diameter tubing with a wall thickness of 049 inch The upright housing tubing shall be swaged to an inside diameter of 640, + 005, - 000 inch at the upper (non-threaded) end The telescoping tube shall be 5/8 inch outside diameter tubing with a wall thickness of 049 inch As an alternate, the outside diameter of the telescoping tube may be 3/4 inch and the upright housing need not be swaged The upright housing tube shall be 32 1/2 + 1/8 inches long, the telescoping tube shall be 26 + 1/8 inches long The tubes shall move smoothly throughout their full length without binding The telescoping tube shall be readily located at any height, by means of a sturdy friction-type mechanism such as friction springs The friction-type mechanism shall not include a thumbscrew, or other manually-operated part which could become detached and lost Friction-locking device shall be capable of having repeated height adjustments of the lamp head made without evidence of failure or necessity of readjustment A stop shall be provided near the lower end of the telescoping tube to prevent inadvertent removal of the telescoping tube from the upright housing The upright housing shall be securely attached to the base, so that there is no rocking of the upright with respect to the base A cushioned stop shall be provided in the lower end of the upright housing to ensure that at least 7/8 inch of the telescoping tubing shall extend out at all times

3 4 3 Flexible arm The flexible arm shall be made of electric-welded, flexible, steel tubing and shall be of conventional design The tubing shall be 5/8 inch outer diameter, with standard commercial tolerance, and 10 3/4 + 1/8 inches long The flexible arm shall be attached to the upper end of the telescoping tube and to the reflector by means of suitable, removable, threaded adapters The overall length of the flexible arm with adapters, shall be 12 + 1/8 inches. The bottom of the flexible arm assembly shall be properly bushed to receive the electric cord The flexible arm shall be capable of bending through at least 120° of arc

3 4 4 Socket and switch The socket (lampholder) shall be 660 watts, 250 volts medium base, with brass or ceramic socket body and cap, and shall conform to the requirement of UL 153. The socket shall be securely fastened inside the reflector so that it shall not become loose or rotate. A suitable single-pole switch with non-metallic control and casing, conforming to the requirements of UL 153, shall be integral to the socket or shall be mounted in a convenient location in the reflector body. The switch shall be of the key, turn-knob, or push-button (but not push-through) type.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

3.4.5 Reflector. The reflector shall be securely attached to the end of the flexible arm so that it shall not turn with respect to the arm. The reflector shall be constructed and vented so that the temperature of the outer surface shall not exceed 50°C (122°F) at any point when tested as required in paragraph 4 4.5 The reflector shall be of single or double shell construction The reflector shall be not more than 8 inches wide nor more than 7-1/2 inches long The bulb shall not extend to within 3/4 inch of the plane of the open end of the reflector The light shall provide illumination of at least 200 foot candles in the center, and at least 100 foot candles at the periphery, of a two-foot diameter circle at a distance of two feet from the front of the reflector when tested as specified in 4.4 4 with an incandescent bulb of 100 watts or less. The interior of the reflector shall be marked legibly and permanently, "Use _____ watt or less", with the wattage that enables the end item to meet both illumination and temperature requirements specified herein No direct light, or light reflected from the inner surface of the reflector, shall be emitted through vents or apertures in the back or side of the reflector.

3 4 6 Lamp cord Cord shall be flexible, 3-conductor rubber-covered, No 18 American Wire gage or heavier, 11 feet in length measured from plug to the flexible arm port, and equipped with a Hospital Grade, grounded, three-prong plug as specified in UL 498, which is compatible with the receptacle specified in ANSI C73 11. The grounding-wire connector shall be secured to the frame or switch mechanism enclosure by a screw, that need not be removed during any servicing operation not involving the power supply cord, or by other equivalent means. Ordinary solder alone shall not be used for securing the grounding conductor Suitable strain relief shall be provided in accordance with UL 544

3 5 Dielectric strength The light shall have no electrical breakdown when tested as specified in 4 4.2.

3 6 Threaded connections The threads of the upper end of the upright housing, the telescoping tubing the flexible arm and adapters and the nut or ring for fastening the reflector to the flexible arm shall be properly made, well cut and of adequate length, so as to assure secure connections which may be readily disconnected and reassembled for maintenance purposes Each threaded connection shall consist of not less than 4 complete threads The upright housing shall be fastened securely to the base so that it shall not rock with respect to the base. The light shall exhibit no sign of wear or looseness after being disassembled and reassembled at least five times (see 4 4 3)

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the current version before use.

3.6.1 Cast iron base connection The upright housing shall be threaded for 1-1/2 inches at the lower end with 20 or more threads per inch. A lock nut, at least 3/16 inch thick with 1-1/8 inches outer diameter, shall be threaded on the upright housing above the base cover The base casting, which shall be at least 11/16 inch thick for an area at least 2 inches in diameter, coaxial with the hole, shall be threaded A nut, at least 1/2 inch thick shall be fastened onto the lower end of the upright A locknut or lockwasher shall be used As an alternate, the upright housing shall pass through an unthreaded hole in the casting, which shall be at least 11/16 inch thick for an area of at least 2 inches in diameter coaxial with the hole. A lock washer and a flanged nut, at least 2 inches in diameter and 1/2 inch thick, shall be threaded onto the bottom of the upright housing, locking it securely in place

3 6.2 Pressed steel base connection The upright housing shall be threaded for 7/8 inch at the lower end with at least 20 threads per inch A lock nut 3/16 inch thick with 1-1/8 inch outer diameter shall be threaded on the upright tubing above the base. The base, which shall be at least 3/16 inch thick, shall be threaded. A nut at least 1/2 inch thick shall be fastened on the lower end of upright A locknut or lockwasher shall be used

3 7 Finish All metal parts shall be thoroughly cleaned to remove all traces of grease, oil dirt, and rust Carbon-steel parts shall be chromium plated over nickel Chrome plate shall be as specified in QQ-C-320 for class 1, type I Nickel plate shall be as specified in QQ-N-290, for class 1, grade F (over steel), or grade E (over brass) Spun aluminum reflector shall be brightened and given an anodic coating to produce a uniform reflective surface in accordance with the provisions of MIL-A-8625 As an alternate, corrosion-resisting steel hardware shall be acceptable

3 8 Identification marking Each list shall be permanently and legibly marked, with the name or registered trademark of the manufacturer

3 9 Service data The contractor shall furnish, with each unit, two copies of service data containing complete instructions for assembly operation maintenance and a list of component parts All information contained in the manual shall be indexed so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identification Identical or interchangeable replacement parts shall be identified by a single part number only

3 10 Workmanship Workmanship shall be first class throughout. Lights shall be free from defects which detract from their appearance or which may impair their serviceability

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1 1 Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4 1 2 Inspection Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item..

4.1 2 1 Classification of inspection The inspection requirements specified herein are classified as follows

a First article inspection.
b Quality conformance inspection.

4 1 2 1 1 First article inspection First article inspection shall be performed on the light when a first article sample is required (see 3 1) This inspection shall include examination of 4 3 and tests of 4 4.

4 1 2 1.2 Quality conformance inspection Quality conformance inspection shall include the examination of 4.3 and tests of 4 4

4 1 3 Certificates of quality Certificates of quality, supplied by the manufacturer of the metals may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the currentversion before use.

4.2 Sampling.

4 2 1 For examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I

TABLE I. Sampling for examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1 0
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4 2.2 For tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II Acceptance level shall be zero for all sample sizes

TABLE II Sampling for tests

Component. Unit of Product	Characteristic	Requirement	Test Procedure	Inspection level
Light	Stability (Tilt)	3 4.1	4 4.1	1/
	Dielectric strength	3 5	4 4.2	100%
	Ease of assembly	3 6	4 4 3	S-1
Reflector	Illumination intensity	3 4.5	4 4.4	1/
	Temperature	3 4 5	4 4 5	1/

1/ one unit from each lot

4 3 Examination The lights shall be examined to determine compliance with all requirements contained in this specification

4.3 1 Classification of defects Examination shall be conducted in accordance with the classification of defects as shown in table III Examination shall not be restricted to the classified possible defects listed in table III

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Component part missing
102	Reflector fails to retain any set position
103	Telescoping tube fails to move easily within upright or fails to retain any set position
104	Limits of height adjustment measured from top of telescoping tube not as specified.
105	Component parts in assembled light not free of cracks, tears, and ruptures
106	Finished surfaces not uniform or free of unplated or corroded areas.
107	Assembled light not stable when resting on a flat surface
108	Upright housing rocks with respect to base.
109	Ground wire not adquately fastened to the frame or enclosure.
110	Assembly has sharpness, burrs, or slivers.
111	Assembled light fails to operate properly when connected to electrical power.
Minor	
201	Identification marking omitted, illegible, or not as specified.
202	Light cord not of gage or length specified.
203	Assembled light not free of dirt, grease, oil, chips and foreign matter.

4.3.2 Dimensional examination. The lights shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.3.3 Examination of preparation for delivery. An examination shall be made to determine compliance with the packaging, packing and marking requirements of section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with AQL of 4.0 expressed in terms of percent defective.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

TABLE IV Examination of preparation for delivery

Examine	Defect
Contents	Not as specified
Container	Not as specified
Marking	Omitted, incorrect illegible, improper size, location sequence or method of application
Packaging and Packing Materials	Component missing or damaged.
Workmanship	Bulging or distortion of containers

4.3 3 1 Examination of closure waterproofing and banding of container When the shipping container is required to comply with PPP-B-636, the examination for defects in closure, waterproofing or banding shall be in accordance with the appendix of the box specification

4 4 Tests Tests shall be conducted to determine compliance with specification requirements

4 4 1 Stability test The light shall be adjusted in height so that the top of the telescoping tube is at its highest position The flexible arm shall be bent so that a line connecting the reflector end and the upper end of the telescoping tube forms a 45° angle with the horizontal The light shall be placed upon a 20° inclined plane which shall slope down in the direction of the arm (if a three-pad base is used, the downward slope shall be midway between two pads)

4 4 2 Dielectric strength test The light shall be subjected to a gradually increasing potential of twice the rated voltage plus 1000 volts which shall be maintained at this value for a period of one minute in accordance with UL 544

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the currentversion before use.

4.4.3 Assembly test. The light shall be disassembled into its shipping configuration; as a minimum, the base, the upright housing, and the flexible arm shall be separate from one-another. The light shall be fully assembled and disassembled five times

4 4.4 Illumination intensity test A previously unused incandescent bulb, of the type and wattage specified by the manufacturer, shall be inserted into the light socket and the light shall be energized with 115 volts, 60 hertz power, in a room in which no other light is present An illumination meter which measures illumination in foot candles within an accuracy of ± 5 percent shall be used to make the measurements indicated. Measurements shall be taken at points within a 24 inch diameter circle located 24 inches from the front edge of the reflector and centered directly in front of the reflector With the sensor at the center of the circle, a measurement of illumination intensity, in foot candles, shall be taken. The illumination shall also be measured at four equidistant points on the circumference of the 24 inch circle. Any measurement which fails to meet the criteria of 3 4.5 shall be classified as a defect

4 4.5 Temperature test This test shall be conducted using the identical light bulb employed to pass the illumination intensity test (4 4 4) The light shall be left on for at least 30 minutes with the illumination directed straight down. After 30 minutes, the temperature of the outside of the reflector shall be measured with a thermocouple or an optical pyrometer with an accuracy of ± 5 percent. The temperature at the hottest point on the outside of the reflector shall not exceed the maximum permitted in 3.4.5.

5 PREPARATION FOR DELIVERY

5 1 Packaging Packaging shall be level A or C as specified (see 6.2)

5.1.1 Level A

5.1 1 1 Unit package Each light shall be disassembled as far as is practical The reflector and tubing shall be wrapped in flexible corrugated fiberboard or other suitable material in such a manner as to prevent damage or marring The light with components shall be further packaged in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic The box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of or damage to the contents. Closure of box shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

5 1.2 Level C Lights shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5 2 Packing Packing shall be level A, B or C as specified (see 6 2)

5 2 1 Level A Four lights shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3, PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification

5 2 1 1 Case liner Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5 2 2 Level B Four lights shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3, PPP-B-601, domestic type PPP-B-621, class 1, or PPP-B-636, type CF class domestic Closure shall be as specified in the appendix of the applicable box specification.

5 2 3 Level C Lights shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order

5.3 Marking

5 3 1 Civil agencies Marking for shipment shall be in accordance with Fed Std No 123

5 3.2 Military activities Each unit package and exterior container shall be marked as specified in MIL-STD-129

6 NOTES

6.1 Intended use. The light is intended for use in clinics and hospitals.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a Title, number and date of this specification.
b Selection of applicable levels of packaging and packing (see 5.1 and 5 2)
c When a first article is required (see 3.1, 4.1.2.1 and 6.3)

6.3 First article. When a first article is required, it shall be tested and approved under the appropirate provisions of 7-104.55 of the Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedures for delivery and other specific instructions for inspection and approval of the first article.

6 4 This specification covers the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-706-5100	LIGHT, FLOOR, Operating and Examining, 115 volt

6 4 Drawing DPSC Drawing No. 24180, which was formerly a part of this specification, has been deleted to remove restrictions on the design and to permit greater design flexibility

MILITARY CUSTODIANS
Army - MD
Navy - MS
Air Force - 03

Preparing activity
DSA-DM

Civil Agency Coordinating Acitvity:
VA-DMS
COM-NBS
GSA-FSS

Project No. 6530-1362

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superinetndent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 35 cents each.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay or will otherwise enhance use of the document DoD contractors, government activities or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
W-L-316B LIGHT, FLOOR

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE:
A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?
☐ YES ☐ NO (If "Yes" in what way?

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974— 603 680 6126 2 1

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PEPSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.