A-A-0051596 (DM)
4 August 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers dental ultrasonic prophylaxis unit handpiece inserts suitable for replacement inserts, for prophylaxis unit models 660, 2001, and 2002, as specified below:

Style I - No. 3, Flat-Hoe Chisel Type
Style II - No. 1000, Curette Type

Salient characteristics:

The ultrasonic prophylaxis unit handpiece inserts, style I and style II, shall be suitable for removing calculus and scale from tooth enamel.

Style I. The insert shall be used for removing supragingival calculus from labial and lingual surfaces of teeth and removing surface stain.

Style II. The insert shall be suitable for use on all supragingival and shallow subgingival tooth surfaces.

The inserts shall include an anti-rotation key, to prevent rotation of the tip. The operative tip shall be in line with the central axis of the handle. The flow through the groove in the tip, working in conjunction with the metal eyelet at the juncture of the tip and nozzle shall assure maximum cavitation of water at the tip with a minimum formation of droplets at the juncture of the tip and nozzle.

The complete insert shall be autoclavable as specified.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Materials and dimensions. This item description shall be in accordance with GG-I-526, as specified herein.

Tip. The tip shall be fabricated of AISI type 420 stainless steel, as specified in GG-I-526.

Connecting body. Style I shall be fabricated of Monel 400; style II shall be fabricated of AISI 17-4 PH stainless steel, as specified in GG-I-526.

Nozzle. The nozzle shall be fabricated of polysulfone.

Stack. The stack shall be fabricated of Permanickel Alloy 300 strip. One strip shall be 0.0095-0.0105 inch thick and 0.178-0.187 inch wide. The thickness of the stack shall be 0.159-0.168 inch thick.

Hardness. The operative tip shall have a Rockwell hardness of Rc 40 to 45.

Finish. The operative tips shall have a mirror or bright finish.

Workmanship. The nozzle shall show no evidence of cracking, holes, or pitting, stack shall be free from metal slivers and burrs which will detract from its appearance or impair its serviceability.

Identification. The inserts shall be identified with the name of the manufacturer.

Unit. Each (EA). One insert of one style only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one handpiece insert of each style. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Autoclavability. The inserts shall be cleaned in accordance with the instructions supplied by the manufacturer. The inserts shall then be steam autoclaved at 121°C, 15 psi for 30 minutes for three cycles. After the inserts are removed from the sterilizer, the entire operative end shall be examined for evidence of corrosion and tested for hardness.

Hardness. Testing for hardness shall be conducted in accordance with American Society for Testing and Materials (ASTM) Designation E-18 for Rockwell hardness.

Contractor certification. The contractor shall certify that the product offered meets the the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Stock Number	Item Identification
6520-01-043-5022	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT, No. 3 Flat-Hoe Chisel Type
6520-01-137-8471	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT, No. 1000, Curette Type

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

05867

PROJECT NO. 6520-1969

Location: NCR - Disk 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEPARTMENT OF THE ARMY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12062 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE ARMY

Commander
Defense Personnel Support Center
Directorate of Medical Materiel
ATTN: DPSC-RST (9-3-F)
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
-0051596 (DM)	INSERT, HANDPIECE, DENTAL ULTRASONIC PROPHYLAXIS UNIT

NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.