A-A-51576A
19 March 1990
SUPERSEDING
A-A-0051576 (DM)
28 July 1986

COMMERCIAL ITEM DESCRIPTION

PREVENTIVE DENTISTRY PASTE (FLUORIDATED)

The General Services Administration has authorized the use of this Commerical Item Description.

Salient characteristics:

This commercial item description covers a fluoridated preventive dentistry paste suitable for use in dental prophylaxis procedures.

The preventive dentistry paste shall be of the following types, as specified below:

Type I - Fine Grit Paste
Type II - Medium Grit Paste

The paste shall be a fluoridated, non-splattering preventive dentistry paste. The paste shall be capable of removing dental plaque, materia alba, and extrinsic stain from tooth structure with insignificant splattering and without undue abrasion, while imparting a polished luster to the teeth. The paste shall have a pleasant orange, cherry, mint, or tropical fruit flavor.

Type I paste (fine grit) shall be gold in color and type II paste (medium grit) shall be blue in color. The color additive shall be approved and certified by the Food and Drug Administration under the Federal Food, Drug and Cosmetic Act.

Physical characteristics. The paste shall contain sodium fluoride with a free fluoride ion content of 1.10 to 1.40% by weight and a cellulose-silicate-glycerin anti-splatter aqueous binder of suitable consistency.

Type I paste shall contain 25 to 30% by weight of diatomaceous silica with a mean particle size of 8 to 10 microns (0.00032 to 0.00040 inch). Type I paste shall have a maximum Relative Enamel Abrasion (REA) value of 8, a minimum Cleaning Ratio of 162, and a minimum Polish Ratio of 12, when tested by the in vitro methods developed by the Oral Health Research Institute of Indiana University (Indianapolis) School of Dentistry.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Type II paste shall contain 50 to 55% by weight of pumice with 72% of the particles smaller than 43 microns (0.0017 inch). Type II paste shall have a maximum REA value of 22, a minimum Cleaning Ratio of 205, and a minimum Polish Ratio of 9, when tested by the methods referenced above.

Shelf life. The paste shall have a minimum shelf life of 24 months from date of manufacture. Not less than 20 months of the expiration dating period shall remain at the time of delivery to the Government.

The paste shall be packaged in a 200 2-gram unit-dose, clear plastic cups sealed with a removable laminated metallized cover to prevent evaporation. The minimum weight of paste in the 200 cups shall be 380 grams for Type I, and 400 grams for Type II. The cup shall be designed to fit into a prophy finger ring holder. A holder shall be included with each unit of issue.

Instructions. Complete and accurate instructions for use of the paste shall be included with each unit of issue.

Workmanship. The paste shall be free from defects which detract from its appearance or impair its serviceability, such as separation, settling, and color non-uniformity.

Unit. Package (PG). One package containing two hundred 2-gram unit dose cups of paste, as specified, constitutes one unit. A prophy finger ring holder for the cups and instructions for use of the paste shall be included in the package. All components shall be packaged in a suitable box. The box shall be marked with the lot or control number and the expiration date.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package of prophylaxis paste. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, and sampling for tests shall be conducted at inspection level S-1.

Examination. The prophylaxis paste shall be examined for defects including, but not limited to, those listed below:

Defects.

Paste not homogeneous or shows separation or settling.
Paste does not have proper consistency.
Color of paste not as specified.
Paste not pleasantly flavored.
Immediate container not a plastic cup as specified.
Package does not contain 200 cups of paste.
Package not marked as specified.
Instructions for use of the paste missing, incorrect, illegible, or incomplete.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-01-219-3867	PASTE, PREVENTIVE DENTISTRY, Fine Grit Silica, Fluoridated, 2 Gram, 200s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

6520-01-219-3873 PASTE, PREVENTIVE DENTISTRY, Medium Grit Pumice, Fluoridated, 2 Gram, 200s

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

59107
59108

PROJECT NO. 6520-2352
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51576A	Preventive Dentistry Paste (Fluoridated)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ASTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.