A-A-0051587 (DM)
4 August 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

CHISEL, DENTAL, Rhodes No. 36 and 37

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This commercial item description covers a double-ended back action periodontal chisel of Rhodes design.

Salient characteristics:

The chisel shall be in accordance with the applicable requirements of GG-I-526 with additional requirements specified herein.

Material. The operative ends shall be class 4, Type 440A, 440B, or 440C stainless steel. Handle shall be chromium plated brass. Brass shall conform to copper alloy No. 260 or 330 as specified in ASTM B135 for chemical composition only.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A, C54 to C58 for type 440B, and C56 to C60 for type 440C when tested in accordance with ASTM E-18 (Rockwell) or E-92 (Vickers or Diamond pyramid).

Style, design and dimensions. The chisel shall be a double-ended back action periodontal chisel No. 36 and 37 of Rhodes design. The chisel shall be suitable for use in periodontal therapeutic procedures. The chisel shall permit access to deep regions of the oral cavity. The chisel shall be of three piece construction. Handle shall be hollow so as to provide lightness and thereby facilitate tactile sensibility. The operative ends shall be permanently secured to the handle. The interface between the operative ends and the handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified herein.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

The style, design and dimensions of the chisel shall conform to DPSC drawing 24532.

Finish. The operative ends and shanks shall have a smooth mirror or bright finish, except for the surfaces (two side surfaces, one lead edge surface and the inside face) that form the three cutting edges which shall have a finely honed finish. The operative end cutting edges shall be sharp and shall be free of defects such as rough grind marks, cracks, gouges, burrs, feathers, or other imperfections when viewed under four power (4X) magnification. Handles shall have a mirror or bright finish, except for the serrated or knurled finger grip portions. Finger grip portions of handles shall provide adequate finger grip during use.

Corrosion resistance. The copper sulfate test shall apply to the operative ends. The boil test shall apply to the entire chisel.

Identification marking. In addition to the requirements of GG-I-526, each chisel shall be marked with chisel identification "C 36/37". All marking shall be on the smooth portion of the handle and shall be permanent and legible. Each chisel shall be marked with the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless", when applicable, shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

Workmanship. The chisel shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One chisel, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one instrument. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Chemical composition. In absence of certificate of quality for the raw material, a sufficient quantity of each lot of metal used in the construction of the chisel may be selected for composition analysis in lieu of sample chisels, provided lot identity has been maintained and can be demonstrated to the Government. Analysis shall be in accordance with methods 111 or 112 of FED-STD-151.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of one second. The instrument shall be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double ended instrument.

Contractor certification. The contractor shall certify that the product offered meets the the salient characteristics of this description and conform to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-01-199-3290	CHISEL, DENTAL, Rhodes 36/37, Back-Action, Corrosion-Resisting Steel, Double-ended: 6-9/16 inch (167mm) long.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS

59091

PROJECT NO. 6520-1967

Location: NCR - Disk 2

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44778

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.