A-A-51587A
15 February 1988
SUPERSEDING
A-A-0051587 (DM)
4 August 1986

COMMERCIAL ITEM DESCRIPTION

CHISEL, DENTAL, (Rhodes No. 36 and 37)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a double-ended back action periodontal chisel of Rhodes design.

Salient characteristics:

The chisel shall be in accordance with the applicable requirements of Federal Specification GG-I-526 with additional requirements specified herein.

Material. The operative ends shall be class 4, Type 440A, 440B, or 440C stainless steel. Handle shall be chromium plated brass. Brass shall conform to copper alloy No. 260 or 330 as specified in ASTM B135 for chemical composition only.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A, C54 to C58 for type 440B, and C56 to C60 for type 440C when tested in accordance with ASTM E-18 (Rockwell) or E-92 (Vickers or Diamond pyramid). Conversion of Vickers Hardness to Rockwell Hardness values shall be in accordance with ASTM E-140.

Style, design and dimensions. The chisel shall be a double-ended back action periodontal chisel No. 36 and 37 of Rhodes design. The chisel shall be suitable for use in periodontal therapeutic procedures. The chisel shall permit access to deep regions of the oral cavity. The chisel shall be of three piece construction. Handle shall be hollow so as to provide lightness and thereby facilitate tactile sensibility. The operative ends shall be permanently secured to the handle. The interface between the operative ends and the handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified herein.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

The style, design and dimensions of the chisel shall conform to DPSC drawing 24532.

Finish. The operative ends and shanks shall have a smooth mirror or bright finish, except for the surfaces (two side surfaces, one lead edge surface and the inside face) that form the three cutting edges which shall have a finely honed finish. The operative end cutting edges shall be sharp and shall be free of defects such as rough grind marks, cracks, gouges, burrs, feathers, or other imperfections when viewed under four power (4X) magnification. Handles shall have a mirror or bright finish, except for the serrated or knurled finger grip portions. Finger grip portions of handles shall provide adequate finger grip during use.

Corrosion resistance. The copper sulfate test of Federal Specification GG-I-526 shall apply to the operative ends. The boil test of Federal Specification GG-I-526 shall apply to the entire chisel.

Identification marking. In addition to the requirements of Federal Specification GG-I-526, each chisel shall be marked with chisel identification "C 36/37". All marking shall be on the smooth portion of the handle and shall be permanent and legible. Each chisel shall be marked with the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless", when applicable, shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

Workmanship. The chisel shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One chisel, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one instrument. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Chemical composition. In the absence of certificates of quality for the raw material, a sufficient quantity of each lot of metal used in the construction of the chisel may be selected for composition analysis in lieu of sample chisels, provided lot identity has been maintained and can be demonstrated to the Government. Analysis shall be in accordance with methods 111 or 112 of FED-STD-151.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of one second. The instrument shall be examined for movement, looseness, or separation of the operative end from the handle. The test shall be repeated for each operative end of the double ended instrument.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-199-3290	CHISEL, DENTAL, Rhodes 36/37, Back-Action, Corrosion-Resisting Steel, Double-ended: 6-9/16 inch (167mm) long.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITES:
VA-OSS
PHS
FDA-MPQAS

59091

PROJECT NO. 6520-2064

Location: NCR - Disk #2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON, D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51587A	CHISEL, DENTAL, (Rhodes No. 36 and 37)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *Last First MI* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.