A-A-51590
12 August 1986

COMMERCIAL ITEM DESCRIPTION

PADDLE, PNEUMATIC COMPRESSION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a pneumatic compression paddle suitable for use in radiographic studies.

Salient characteristics:

The pneumatic compression paddle shall be for use in radiographic studies of the gastrointestinal tract.

The paddle shall be constructed of a suitable plexiglas material with an imbedded metal ring for quick fluoroscopic localization.

The paddle shall be equipped with an inflatable rubber bladder with an inflating bulb, and a thumb needle valve to secure bladder in inflated state.

The paddle shall have a non-skid bottom to prevent slipping out of position.

Bladder, inflating bulb, and needle valve shall be free of leakage.

The paddle shall be approximately 22 inches (56.0 cm) long, 7 inches (17.8 cm) wide, and 1/2 inch (1.3 cm) thick.

The paddle shall be free of burrs, cracks, flash, and sharp or ragged edges.

Each paddle shall be furnished with a list of replaceable component parts.

Workmanship. The pneumatic compression paddle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One pneumatic compression paddle, as specified, constitutes one unit.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-139-9435

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS

PGC 52871

Project No. 6525-1197

Location: CID/PADDLE/H13-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.