A-A-0051602(DM)
8 August 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

CANISTER, ASSEMBLY, SUCTION, SURGICAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a plastic suction canister assembly specifically designed to collect and contain fluids during surgical procedures.

Salient characteristics:

The suction canister assembly shall consist of a 1500 ml disposable plastic inner canister which fits into a compatible autoclavable outer canister. On reusable outer container with lid shall be provided with 100 inner canisters.

The outer canister shall be calibrated with markings of at least every 25 ml to the 100 ml level, every 50 ml to the 300 ml level and every 100 ml thereafter.

Both the inner and outer canister walls shall be clear for high visibility of aspirated contents. Inner canisters shall have crack resistant walls. Outer canisters shall have rigid crack resistant walls. The inner canisters shall be disposable through incineration.

Inner canisters shall be of a shape allowing nesting for reduction of storage space. As an alternate, inner canisters shall be collapsible to allow for reduction of storage area. Each inner canister shall be provided with a snap seal lid which prevents leakage and spilling during aspiration disposal. Collapsible items may have an integral lid sealed to the bag.

The canister assembly shall have the following ports.

a. Vacuum port. The vacuum port shall have a built in filter to minimize bacterial spray into the vacuum system.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

b. Patient port. Patient port shall include a right angle swivel connector for elimination of patient tube kinking.

c. Accessory port. Shall be a pour spout designed to allow canister contents to be emptied in an area isolated from patients and staff.

The suction canister assembly shall operate properly using a vacuum of up to 30 inches Hg. The outer shall be autoclavable up to a temperature of 250°F.

Workmanship. The suction canister shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred disposable inner canisters along with one reusable outer canister, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one canister. Sampling for examination shall be in accordance with table I and sampling for tests shall be conducted at inspection level S-1, AQL 1.0.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
Major defects	II	1.0
Total defects (Major and minor)	II	3.0

Examination. The suction canister shall be examined to determine compliance with the salient characteristics specified herein. Examination shall include but not be limited to the classification of defects in table II..

TABLE II Categories and Defects *

Categories	Defects
Major	
101	Suction canister not complete with all required components.
102	Suction canister cracked, broken, or otherwise unusable.
103	Edges and surfaces not free of sharpness or foreign imbedded materiel.
Minor	
201	Suction canister not free of dirt or loose debris.

Tests. Tests shall be conducted to determine compliance with the salient characteristic specified herein. Failure of any test shall be considered a major defect. Testing shall include but not be limited to the following:

1. The suction canister shall be checked for proper operation with a vacuum of 30 inches Hg.

2. The reuseable outer canister shall be autoclaved at a temperature of 250°F and then examined for defects.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-106-7648

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51345

Project No. 6515-3647

Location: CID/CANISTER/H10-10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.