INCH-POUND

A-A-51602B
5 August 1994
SUPERSEDING
A-A-51602A
31 August 1987

COMMERCIAL ITEM DESCRIPTION

CANISTER ASSEMBLY, SUCTION, SURGICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a suction canister assembly for collecting fluids during surgery.

Salient characteristics:

The suction canister assembly shall consist of a 1500 ml disposable plastic inner canister which shall fit into a compatible autoclavable outer canister and a wall bracket specifically designed to hold the outer canister. One reusable outer container with lid and one compatible wall bracket shall be provided with 100 inner canisters. The canister assembly shall meet the requirements of ASTM F960 portions pertaining to suction collection vessel (not pleural and mediastinal drainage), tubing, and connectors as applicable.

The outer canister shall be calibrated with markings of at least every 25 ml to the 100 ml level, every 50 ml to the 300 ml level, and every 100 ml thereafter.

Both the inner and outer canister walls shall be clear for high visibility of aspirated contents. The inner canisters shall be disposable through incineration.

The inner canisters shall be of a shape allowing nesting for reduction of storage space. As an alternate, inner canisters shall be collapsible to allow for reduction of storage area.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Each inner canister shall be provided with a snap seal lid which prevents leakage and spilling during aspiration and disposal. Collapsible items may have an integral lid sealed to the bag.

The canister assembly shall be provided with a shut-off valve to protect suction source and shall have the following ports:

* Vacuum port. The vacuum port shall have a built in filter to minimize bacterial spray into the vacuum system.

* Patient port. The patient port shall include a right angle swivel connector for elimination of patient tube kinking.

* Accessory port. The accessory port shall incorporate a pour spout designed to allow the canister contents to be emptied in an area isolated from patients and staff.

The suction canister assembly shall be capable of being exposed to an implosion vacuum of 690 mm Hg without degradation (implosion vacuum measured in occluded inner canister), and shall be capable of operating while being exposed to vacuums up to 500 mm Hg as measured at or near the patient end of 48 inch long patient tube with catheter end occluded and having 300 to 500 ml of water in the canister.

The outer canister shall be autoclavable up to a temperature of 250°F for at least one sterilization cycle.

Workmanship. The suction canister assembly with the 100 inner canisters shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing one hundred disposable inner canisters along with one reusable outer canister, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 1.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one canister assembly with 100 inner canisters), as specified.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for test purposes shall be one assembled canister assembly (using one inner canister), as specified.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

* All applicable tests in ASTM F960 shall apply

* Test the implosion requirement and operating vacuum requirement. During the latter test, compare the reading of water in the canister with the actual amount of water. A deviation of more than 10 percent is not acceptable.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

NOTE 1: This CID is referenced in the specification covering NSNs 6515-01-167-7287, 6515-01-180-6984, and 6515-01-161-2938, Suction Apparatus, Types I, II, and III respectively. Any exceptions that may become effective to this specification should be noted in the procurement document for the Suction Apparatus, Type I or III (Type II does not include this canister assembly).

NOTE 2: The following National Stock Number is covered by this document.

6515-01-106-7648

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5607

Location: CANS12C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51602B	940805

3. DOCUMENT TITLE

CANISTER ASSEMBLY, SUCTION, SURGICAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER		
a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY		
a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	b. TELEPHONE (Include Area Code) (1) Commercial (215)737-8431	(2) AUTOVON 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete. — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.