A-A-51603A
17 April 1987
SUPERSEDING
A-A-0051603(DM)
8 August 1986

COMMERCIAL ITEM DESCRIPTION

RESUSCITATOR, CARDIOPULMONARY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable resuscitator suitable for emergency cardiopulmonary resuscitation in a variety of locations and during patient transport.

Salient characteristics:

The synchronized cardiopulmonary resuscitator shall be a portable, oxygen powered, automatic unit. The resuscitator shall supply intermittent time cycled, positive pressure ventilation and external cardiac compression via the patient's sternum.

Subassemblies.

The resuscitator shall consist of two main parts: a base/ventilator assembly, and a column/compression arm assembly.

The base/ventilator assembly shall consist of a master valve, a compression on-off valve, a force control knob, a ventilation on-off valve, a ventilation pressure control knob, an applied chest force knob, and a ventilation pressure gauge.

The column/compression arm assembly shall contain a piston which depresses the sternum during resuscitation. The compression arm shall be adjustable to accommodate various sizes. The column arm and the piston arm shall be color coded for accurate volume sternal compression.

Power source.

The resuscitator shall operate from a regulated compressed oxygen source of 40 to 85 psi (2.81 to 6.0 Kg/cm^2).

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Controls.

Compression force. The resuscitator shall have an adjustable compression force range of 0 to 145 pounds (0 to 65.8 Kg) when operated from a supply pressure of 50 psi.

Compression stroke. The compression stroke shall be adjustable from 0 to 3/4 inches (0 to 1.90 cm).

Ventilation pressure. The unit shall have an adjustable ventilation pressure range of 5.0 to 60.0 cm H_2O.

Compression frequency. The compression frequency shall be adjustable from 40 to 120 cycles/minute.

Chest time pressure. The chest time pressure shall be adjustable from 30 percent to 70 percent of compression cycle.

Ventilation duration. The unit shall have an adjustable ventilation period from 0.5 to at least 2.0 seconds.

Components.

The components of the unit shall at least include the following:

a. Carrying/storage case with hook and pile straps for holding backboard.

b. Transport model backboard with straps.

c. Mobile oxygen carrier for "D" size cylinders.

d. DISS pipeline adapter.

Performance.

Ventilation output. The unit shall have a ventilation output of 140 L/minute, minimum at 20 cm H_2O backpressure.

Materials.

Metal parts. Alloy metal parts shall be fabricated of a suitable metal alloy that has been properly hardened and passivated.

Plastic parts. All plastic parts shall be fabricated of materials which are in accordance with USP Class IV, Plastic.

Instructions. Each unit shall include one (1) each user's manual and two (2) each repair manuals.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Workmanship. The resuscitator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One resuscitator, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-160-8371

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

A-A-51603A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4068

PGC 51899

Location: CID/RESUSCITATOR/H10-19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.