A-A-51606
8 August 1986

COMMERCIAL ITEM DESCRIPTION

SKIN CARE KIT, INTRAVENOUS CATHETER

The General Services Administration has authorized the use of this commercial item description.

Salient characteristics:

The skin care kit shall be a sterile and disposable I.V., suitable for use during routine catheter cleansing procedures, facilitating more effective control of hospital-associated catheter injection.

The kit shall consist of the following components:

a) One underwrap, Poly-Back Towel, 13 ± 1/2 by 19 ± 1/2 inches.
b) One pair exam gloves of universal sizing.
c) One povidone-iodine ointment packet, 1.2 grams net weight.
d) One gauze sponge, 4 ± 1/4 by 4 ± 1/4 inches.
e) 4 rayon balls.
f) One plastic basin, 6 oz (177 ml).
g) One povidone-iodine solution packet, 1 oz (29.5 ml) which shall be film-forming, water soluble and non-staining to skin and natural fabrics.
h) One wooden cotton-tipped applicator.

The skin care kit kit shall have not less than 5/6 of its shelf-life remaining before receipt by the Government.

Sterility. The skin care kit shall be sterile when tested in accordance with USP. Sterility assurance level shall be 0.0001 as a minimum.

Workmanship. The skin care kit shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing 50 catheters kits, each kit individually packaged sterile in a peel open type container, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The unit of product for test purposes shall be one kit.

Alternate method. Supplies conducting sterility tests or designating lots in a manner other than that indicated in the USP and MIL-STD-105, shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. Request for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a Government contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Product testing. Testing for sterility of the product after receipt by the Government shall be in accordance with the USP. Sterility Assurance Level shall be 0.00001 as a minimum.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-153-5318	SKIN CARE KIT, INTRAVENOUS CATHETER, 50s

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.

Ordering data. Intermediate/exterior package quantities and labelingg and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA -OSS
PHS

PGC 50922

Project No. 6515-3324

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44945

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.