A-A-0051615(DM)
28 October 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

CLIP, HEMOSTATIC, Small

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers small size hemostatic ligating clips.

Salient characteristics:

General. Shall be small size stainless steel or tantalum hemostatic clips mounted in color-coded cartridges. The clip shall be used for blood vessel occlusion in surgical procedures.

Material. Stainless steel clips shall be fabricated of 316L stainless steel in accordance with ASTM F56 or F138, Grade 2. The tantalum clips shall be fabricated material in accordance with ASTM F560. Cartridges shall be fabricated of a suitable rigid plastic.

Style and design. The clips shall be in accordance with DPSC Drawing 23391.

Clip. The shape of the clips shall resemble that of a staple with its inner surface textured to prevent slippage of the vessel being ligated. The center of the stainless steel clip shall have reduced thickness to aid closure. The ends of the clip shall be serrated to enable pick-up and firm retention in the forceps jaws. The clip shall be 2.5 mm when open and 3.0 mm when closed, as measured from the center of the clip to either end.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Cartridge. The cartridge shall be uniquely color coded to identify and distingish the particular size of the clips from other sizes of clips. The particular color of the cartridge shall match the color of those forceps compatible with small clips.

Compatibility and closure. When subjected to the closure test specified herein, the clips shall exhibit no splitting or fracture.

Sterility. The clips shall be tested for sterility in accordance with procedures contained in the USP. The sterility assurance level shall be 0.000001.

Workmanship. The hemostatic clips shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 36 cartridges, each cartridge containing 6 individually packaged sterile clips in a peel-open type container, capable of maintaining sterility of the contents unless package is damaged or opened, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. At inspection level S-2, dimensional testing shall be at a 1.5% AQL and material and closure testing shall be at a 2.5% AQL. The unit of product for test purposes shall be one hemostatic clip.

Compatibility and closure test. Remove a clip with forceps having the same handle color coding as that for the cartridge. If the forceps fails to engage the clip properly, verify that the forceps is the proper size. If so, the lot is rejected. Close the clip completely. Examine with the naked eye for splits or fractures. A minimum of 6 clips shall be tested.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-156-2603	CLIP, HEMOSTATIC, Stainless Steel, Small, 216 Clips in 36 Cartridges
6515-01-173-8925	CLIP, HEMOSTATIC, Tantalum, Small, 216 Clips in 36 Cartridges

NOTE 2: The items covered by the following National Stock Numbers are compatible with and color coded to match this specification item.

6515-01-167-6670	FORCEPS, HEMOSTATIC CLIP, Applying, Small, Angled 15°, 7-1/2 Inch (19.05 cm)

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 58683

Project No. 6515-3843

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.