A-A-51620
22 August 1986

COMMERCIAL ITEM DESCRIPTION

PAD, HOT-COLD TREATMENT, CIRCULATING

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification O-P-001074

This commercial item description covers a controlled temperature pad unit suitable for hot or cold therapeutic treatment.

Salient characteristics:

The unit shall be a controlled temperature pad for therapeutic treatment involving hot or cold applications either wet or dry.

The pad unit shall consist of:

a. A temperature control unit containing a sealed electric motor combined with a pump; a water reservoir; an electrical heating element in the reservoir; an adjustable thermostat to control the heated water temperature; a removable control knob for the adjustable thermostat; an over-temperature thermostat; a pilot light; and a 3-wire power cord terminated with a 3-prong hospital grade plug.

b. A flexible vinyl plastic pad approximately 12 by 15 inches (30.48 by 38.10 cm) with an appropriate length of twin tubing.

The temperature control unit shall operate on 115 volt, 60 Hz, alternating electrical current.

In heat therapy, the control of the watt heating element shall be such that the temperature control unit will maintain water temperature to within plus or minus 1°F through a range of 85°F - 105°F (29°C - 41°C). An over-temperature thermostat shall be provided which will activate between 108°F and 114°F (42°C and 45.5°C).

The hot-cold treatment pad shall attain operational temperature in less than 10 minutes when filled with water having a temperature of 50°F to 60°F and operated in an ambient temperature of 70°F to 80°F (22°C -28°C).

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

In cold therapy, the pumping element of the temperature control unit shall be used and heat dispersal shall be accomplished through an external cooling unit, or by cooling the circulating tubing in a container of ice cubes.

The pad temperature control unit shall comply with the requirements of Underwriters Laboratories, Inc Standard 544.

Each hot-cold treatment pad shall be furnished with two copies of a technical manual which contain complete instructions for operation and maintenance, and list of component parts.

Workmanship. The hot-cold treatment pad shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One hot-cold treatment pad, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-061-1825

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 50648

Project No. 6530-1512

Location: CID/PAD-H-C-TREATMENT-CIR/H13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.